
Smart Healthcare -- Reality or Hype?

Keynote - International Conference on Artificial Intelligence and Signal Processing (AISP 2022)

Vijayawada, India, 12-14 Feb 2022

Saraju P. Mohanty

University of North Texas, USA.

Email: saraju.mohanty@unt.edu, **Website:** <http://www.smohanty.org>

Outline

- Healthcare → Smart Healthcare
- Smart Healthcare - Characteristics
- Smart Healthcare - Components
- Smart Healthcare - Examples
- Smart Healthcare – Challenges
- Smart Healthcare – Solutions of Challenges
- Smart Healthcare – COVID-19 Perspectives
- Conclusions and Future Directions

Healthcare to Smart Healthcare


Human Body and Health

Human Body


- From an engineering perspective - Human body can be defined as a combination of multi-disciplinary subsystems (electrical, mechanical, chemical ...).

Health

- Human health is a state of complete physical, mental and social well-being.



Traditional Healthcare




- Physical presence needed
- Deals with many stakeholders
- Stakeholders may not interact
- May not be personalized
- Not much active feedback
- Less effective follow-up from physicians

Telemedicine



Telemedicine: The use of telecommunication and information technology to provide clinical health care from a distance.


Electronic Health (eHealth)



eHealth: The use of information technology to improve healthcare services.

Source: W. O. Nijeweme-d'Hollosy, L. van Velsen, M. Huygens and H. Hermens, "Requirements for and Barriers towards Interoperable eHealth Technology in Primary Care," *IEEE Internet Computing*, vol. 19, no. 4, pp. 10-19, July-Aug. 2015.

Connected Health (cHealth)



Source: https://www.slideshare.net/tibisay_hernandez/connected-health-venfinal

cHealth: Connections of the various healthcare stakeholders through Internet to share appropriate data to better serve the patients.


Mobile Health (mHealth)



mHealth: Healthcare supported by *mobile devices* that uses mobile telecommunications and multimedia technologies for the delivery of healthcare services and health information.

Source: H. Zhu, C. K. Wu, C. H. KOO, Y. T. Tsang, Y.Liu, H. R. Chi, and K. F. Tsang, "Smart Healthcare in the Era of Internet-of-Things", *IEEE Consumer Electronics Magazine*, vol. 8, no. 5, pp. 26-30, Sep 2019.


Smart Healthcare (sHealth)



Fitness Trackers



Headband with Embedded Neurosensors



Embedded Skin Patches

Quality and sustainable healthcare with limited resources.

Source: P. Sundaravadivel, E. Kougianos, S. P. Mohanty, and M. Ganapathiraju, "Everything You Wanted to Know about Smart Health Care", *IEEE Consumer Electronics Magazine (MCE)*, Vol. 7, Issue 1, January 2018, pp. 18-28.

Smart Healthcare - Applications



Healthy Living

- Fitness Tracking
- Disease Prevention
- Food monitoring

Home Care

- Mobile health
- Telemedicine
- Self-management
- Assisted Living

Acute Care

- Hospital
- Specialty clinic
- Nursing Home
- Community Hospital

Internet of Medical Things (IoMT)

Frost and Sullivan predicts smart healthcare market value to reach US\$348.5 billion by 2025.

Source: P. Sundaravadivel, E. Kougianos, S. P. Mohanty, and M. Ganapathiraju, "Everything You Wanted to Know about Smart Health Care", *IEEE Consumer Electronics Magazine (MCE)*, Vol. 7, Issue 1, January 2018, pp. 18-28.



Smart Healthcare - Characteristics

What is Smart Healthcare?

Smart Healthcare



Conventional Healthcare

+ Body sensors

+ Information & Communication Technology (ICT)

+ Artificial Intelligence (AI)/ Machine Learning (ML)

+ Smart Technology (BCI, VR, etc.)


Internet of Medical Things (IoMT)

Internet of Health Things (IoHT)

Healthcare Cyber-Physical Systems (H-CPS)

Source: P. Sundaravadivel, E. Kougianos, S. P. Mohanty, and M. Ganapathiraju, "Everything You Wanted to Know about Smart Health Care", *IEEE Consumer Electronics Magazine (MCE)*, Volume 7, Issue 1, January 2018, pp. 18-28.

Smart Healthcare - 4-Layer Architecture



Healthcare Cyber-Physical Systems (H-CPS)

Source: M. Ghamari, B. Janko, R.S. Sherratt, W. Harwin, R. Piechockic, and C. Soltanpur, "A Survey on Wireless Body Area Networks for eHealthcare Systems in Residential Environments", *Sensors*, 2016. 16(6): p. 831.

Wearable Medical Devices (WMDs)



Fitness Trackers



Headband with Embedded Neurosensors




Source: <https://www.empatica.com/embrace2/>

Medical grade smart watch to detect seizure



Insulin Pump

Source: <https://www.webmd.com>






Embedded Skin Patches

Implantable Medical Devices (IMDs)



Pill Camera




Brain Pacemaker

Source: P. Sundaravadivel, E. Kougianos, S. P. Mohanty, and M. Ganapathiraju, "Everything You Wanted to Know about Smart Health Care", IEEE Consumer Electronics Magazine (MCE), Volume 7, Issue 1, January 2018, pp. 18-28.


Collectively:
Implantable and Wearable
Medical Devices (IWMDs)

Implantable MEMS Device



Source: <http://web.mit.edu/cprl/www/research.shtml>

Smart Healthcare – 7Ps



Source: H. Zhu, C. K. Wu, C. H. KOO, Y. T. Tsang, Y.Liu, H. R. Chi, and K. F. Tsang, "Smart Healthcare in the Era of Internet-of-Things", *IEEE Consumer Electronics Magazine*, vol. 8, no. 5, pp. 26-30, Sep 2019.

IoMT Advantages & Limitations

Advantages

Patients/Users

- Real-time interventions in emergency
- Cost reduction
- Reduced morbidity and financial burden due to less follow up visits

Healthcare Service Providers

- Optimal utilization of resources
- Reduced response time in emergency

Manufacturers

- Standardization/compatibility and uniformity of data available
- Capability to sense and communicate health related information to remote location

Limitations

Technical Challenges

- ❖ Security of IoT data - hacking and unauthorized use of IoT
- ❖ Lack of standards and communication protocols
- ❖ Errors in patient data handling
- ❖ Data integration
- ❖ Need for medical expertise
- ❖ Managing device diversity and interoperability
- ❖ Scale, data volume and performance


Market Challenges

- ❖ Physician compliance
- ❖ Data overload on healthcare facility
- ❖ Mobile hesitation
- ❖ Security policy compliance

Source: Y. Shelke and A. Sharma, "Internet of Medical Things", 2016, Aranca, <https://www.aranca.com/knowledge-library/special-reports/ip-research/the-internet-of-medical-things-iomt>, Last Visited 10/18/2017.

Smart Healthcare - Components


Internet of Medical Things (IoMT)



IoMT is a collection of medical sensors, devices, healthcare database, and applications that connected through Internet.

Source: <http://www.icemiller.com/ice-on-fire-insights/publications/the-internet-of-health-things-privacy-and-security/>
Source: <http://internetofthingsagenda.techtarget.com/definition/IoMT-Internet-of-Medical-Things>

Smart Healthcare Sensors



| Types of Sensors | |
|---|----------------------------|
|  | Brain related applications |
|  | Imaging applications |
|  | Heart related applications |
|  | Skin related applications |
|  | Blood related applications |
|  | Ingestible sensors |
|  | Motion Detection |

Smart Healthcare Communication

| Technology | Frequency Band | Data Rate | Range | Transmission Power |
|---|---|---|------------|--------------------|
| Bluetooth 4.0 (LE) | 2.4 GHz | 50–200 Kbps | 30 m | ~10 mW |
| Zigbee | 868 MHz/ 915 MHz/ 2.4 GHz | 20–250 Kbps | 30 m | 30 mW |
| ANT | 2400-2485 MHz | 1 Mbps | Up to 10 m | 0.01–1 mW |
| IEEE 802.15.6 | 2,360-2,400/ 2,400-2,483.5 MHz UWB: 3–10 GHz HBC: 16/27 MHz | NB: 57.5–485.7 Kbps UWB: 0.5–10 Mbps | 1.2 m | 0.1 μ W |
| Medical Implant Communications Service (MICS) | 402-405 MHz | Up to 500 Kbps | 2 m | 25 μ W |

Source: V. Custodio, F.J. Herrera, G. López, and J. I. Moreno, “A Review on Architectures and Communications Technologies for Wearable Health-Monitoring Systems”, *Sensors*, 2012. 12(10): p. 13907-13946.

Electronics Health Record (EHR)

- Electronic Health Record (EHR) is the systematized collection of health information of individuals stored in a digital format.
- Created by various health providers such as hospitals and clinics.

| Meeting | Date | Time |
|----------------------------|-----------|-------|
| 2 month checkup | 5 Mar 09 | 2m.0d |
| 1 month checkup | 5 Feb 09 | 1m.0d |
| Respiration problem | 22 Jan 09 | 17d |
| 10 days chekup | 13 Jan 09 | 8d |
| Control for return at home | 9 Jan 09 | 4d |
| Birth | 5 Jan 09 | 0d |

| Diagnosis | My Diagnosis | Social |
|--------------|-------------------------------------|--------------------------|
| General | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| My Diagnosis | <input type="checkbox"/> | <input type="checkbox"/> |
| Social | <input type="checkbox"/> | <input type="checkbox"/> |

| Notes |
|---|
| Father ask many questions, add 10 minutes to consultation |

Electronic Medical Record (EMR)


Smart Healthcare – AI/ML is Key



- AI Role Includes:
- Automatic diagnosis
 - Disease predication
 - Diet prediction
 - Pandemic projection
 - Automatic prescription

Source: Robert Pearl, "Artificial Intelligence In Healthcare: Separating Reality From Hype", 13 Mar 2018, <https://www.forbes.com/sites/robertpearl/2018/03/13/artificial-intelligence-in-healthcare/?sh=598aa64d1d75>

Brain Computer Interface (BCI)




Source: <http://brainpedia.org/what-is-brain-computer-interface-bci/>



Source: <http://brainpedia.org/brain-computer-interface-allows-paralysis-als-patients-type-much-faster/>

BCI Allows paralysis patients to Type

BCI Allows paralysis patients move a wheelchair



Virtual Reality in Healthcare



Source: <http://medicalfuturist.com/5-ways-medical-vr-is-changing-healthcare/>


For Therapy



Source: <https://touchstoneresearch.com/tag/applied-vr/>


In Surgery

Crowdsourcing for Smart Cities



Smart Healthcare – Specific Examples

What is Stress?



- ❑ Stress is the relationship between a person and a situation, which adversely impacts the happiness and health of the sufferer or physiological reactions.
- ❑ Stress can be divided into **two parts**: stressor and reaction.
- ❑ Stressor is the activity or effect that triggers a change in the physiological parameter values of the human body.
- ❑ Reaction is the deviation of these parameter values from their normal levels.

Stress Monitoring and Control is Needed

Stress is the **body's reaction** to any change that requires an adjustment or response.

Sudden encounter with **stress**

→ Brain floods **body** with chemicals and hormones
(**adrenaline and cortisol**)

- Lack of Energy
- Type 2 Diabetes
- Osteoporosis
- Mental cloudiness (brain fog) and memory problems
- A weakened immune system, leading to more vulnerable to infections



Distress




Eustress

Stress is a Global Issue


- In major global economies - 6 in 10 workers experiencing increased workplace stress.
- In USA: 75% of adults reported experiencing moderate to high levels of stress. 1 out of 75 people may experience panic disorder.
- In Australia: 91% of adults feel stress in at least one important area of their lives.
- In UK: An estimated 442,000 individuals, who worked in 2007/08 believed that they were experiencing work-related stress
- Depression is among the leading causes of disability worldwide. 25% of those with depression world-wide have access to effective treatments → 75% don't have.

Source: <http://www.gostress.com/stress-facts/>

Stress Monitoring & Control – Our Vision



| Sensor | Low Stress | Normal Stress | High Stress |
|---------------------------|------------|---------------|-------------|
| Accelerometer (steps/min) | 0-75 | 75-100 | 101-200 |
| Humidity (RH%) | 27-65 | 66-91 | 91-120 |
| Temperature °F | 98-100 | 90-97 | 80-90 |



Source: L. Rachakonda, S. P. Mohanty, E. Kougianos, and P. Sundaravadivel, "Stress-Lysis: A DNN-Integrated Edge Device for Stress Level Detection in the IoMT", *IEEE Transactions on Consumer Electronics (TCE)*, Vol 65, No 4, Nov 2019, pp. 474--483.

Consumer Electronics Devices – Can Provide Data for Stress Detection

| Brand | Device | Signals | RTI | Ambulant |
|----------|-----------------|-----------------------------|-----|----------|
| Empatica | E4 wristband | PPG, GSR, HR, ACC, ST | Yes | Yes |
| Garmin | Vivosmart | HR, HRV, ACC | Yes | Yes |
| Zephyr | BioHarness 3.0 | HR, HRV, GSR, ACC, ST | Yes | Yes |
| iMotions | Shimmer 3+ GSR | GSR, PPG | Yes | No |
| BIOPAC | Mobita Wearable | ECG, EEG, EGG, EMG, and EOG | Yes | No |

GSR = Galvanic Skin Response, HR = Heart Rate, ACC = Acceleration, ST = Skin Temperature, HRV = Heart Rate Variability, PPG = Photoplethysmograph, RTI = Real Time Implementation



Source: R. K. Nath, H. Thapliyal, A. Caban-Holt, and S. P. Mohanty, "Machine Learning Based Solutions for Real-Time Stress Monitoring", *IEEE Consumer Electronics Magazine (MCE)*, Vol. 9, No. 5, September 2020, pp. 34--41.

Consumer Electronics Sleep Trackers

| Consumer Products | Approach | Features | Drawbacks |
|-----------------------|--------------|--|--|
| Fitbit [34] | Wearable | Heart rate monitor, sleep stages monitor. Has techniques to improve the sleep score. | Relationship between stress and sleep is not discussed. |
| SleepScore Max [36] | Non-wearable | Invisible radio wave sleep tracking | Does not manage stress with sleep. |
| Nokia Sleep [38] | Non-wearable | Uses Ballistocardiography sensor | Does not explain the relationship with stress with sleep. |
| Xiaomi Mi Band 3 [31] | Wearable | Pulse Monitor | No information on importance of quality sleep. |
| Eversleep [32] | wearable | Snoring and breathing interruptions | No explanation on the relationship between stress and sleep. |
| Beddit [35] | Non-wearable | Monitors snoring | Doesn't consider other possible features. |
| Eight [37] | Non-Wearable | Humidity, temperature, heartbeat, breathing rate | No data on how it is important to have a good sleep. |
| Dreem [33] | Wearable | Simulates slow brain waves | It doesn't consider other features; Does not manage stress with sleep. |
| Muse [26] | Wearable | Simulates brain waves | No understanding of the importance of quality sleep. |

Source: L. Rachakonda, A. K. Bapatla, S. P. Mohanty, and E. Kougianos, "SaYoPillow: A Blockchain-Enabled, Privacy-Assured Framework for Stress Detection, Prediction and Control Considering Sleeping Habits in the IoMT", *arXiv Computer Science*, arXiv:2007.07377, July 2020, 38-pages.


Stress-Lysis: From Physiological Signals



Stress-Lysis - DNN has been trained with a total of 26,000 samples per dataset and has accuracy upto 99.7%.

Source: L. Rachakonda, S. P. Mohanty, E. Kougianos, and P. Sundaravadivel, "Stress-Lysis: A DNN-Integrated Edge Device for Stress Level Detection in the IoMT", *IEEE Transactions on Consumer Electronics (TCE)*, Vol 65, No 4, Nov 2019, pp. 474--483.


Stress-Lysis: Experiments



Stress-Lysis - DNN has been trained with a total of 26,000 samples per dataset and has accuracy upto 99.7%.

Source: L. Rachakonda, S. P. Mohanty, E. Kougianos, and P. Sundaravadivel, "Stress-Lysis: A DNN-Integrated Edge Device for Stress Level Detection in the IoMT", *IEEE Transactions on Consumer Electronics (TCE)*, Vol 65, No 4, Nov 2019, pp. 474--483.

iMirror: Our Smart Mirror for Stress Detection from Facial Features



iMirror: Stress detection accuracy - 97%.

Source: L. Rachakonda, P. Rajkumar, **S. P. Mohanty**, and E. Kougianos, "iMirror: A Smart Mirror for Stress Detection in the IoMT Framework for Advancements in Smart Cities", *Proceedings of the 6th IEEE Smart Cities Conference (ISC2)*, 2020.


Smart-Yoga Pillow (SaYoPillow) - Sleeping Pattern

Person On Pillow:

Physiological Sensor Data Monitoring Starts



Period 1. Lying on bed but not Sleeping



Period 2: Trying to Sleep

Period 3: Drift from Wakefulness to Sleep



Period 4: Deep Sleep

Person Off Pillow:

Physiological Sensor Data Monitoring Ends



Period 5: Awake Person




Transitions of a person drifting into non-rapid eye movement (NREM) followed by rapid eye movement (REM) to Awake State.




Source: L. Rachakonda, A. K. Bapatla, S. P. Mohanty, and E. Kougianos, "SaYoPillow: Blockchain-Integrated Privacy-Assured IoMT Framework for Stress Management Considering Sleeping Habits", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. 67, No. 1, Feb 2021, pp. 20-29.

SaYoPillow – Stress Analysis Approach



SaYoPillow – Uses deep learning for 96% accuracy with blockchain based security features



Source: L. Rachakonda, A. K. Bapatla, S. P. Mohanty, and E. Kougianos, "SaYoPillow: Blockchain-Integrated Privacy-Assured IoMT Framework for Stress Management Considering Sleeping Habits", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. 67, No. 1, Feb 2021, pp. 20-29.


Stress Control by iFeliz: Our Proposed System

Generate workout plan, meal plan, sleep schedule, display stress relief paintings, play music in the background, suggest videos to play, quick 2 min breathe exercise, display positive and inspirational quotes, nearby therapy dog's location, automatic slide show of photos from gallery.

Physical exercise, yoga, meditation- heavy breathing, specific music, shower, Massage appointment, Nap, pet time.


Long-Term Advice

Short-Term Advice



Source: L. Rachakonda, S. P. Mohanty, and E. Kougianos, "iFeliz: An Approach to Control Stress in the Midst of the Global Pandemic and Beyond for Smart Cities using the IoMT", in *Proc. of IEEE Smart Cities Conference (ISC2)*, 2020.


iFeliz: Stress Control Approaches



iFeliz - 15 Features, Stress Detection, Stress Control, Accuracy - 97%.

Source: L. Rachakonda, S. P. Mohanty, and E. Kougianos, "iFeliz: An Approach to Control Stress in the Midst of the Global Pandemic and Beyond for Smart Cities using the IoMT", in *Proc. of IEEE Smart Cities Conference (ISC2)*, 2020.

Automatic Food Intake Monitoring and Diet Management is Important



Imbalance Diet is a Global Issue


- Imbalanced diet can be either more or fewer of certain nutrients than the body needs.
- In 2017, 11 million deaths and 255 million disability-adjusted life-years (DALYs) were attributable to dietary risk factors.
- Eating wrong type of food is potential cause of a dietary imbalance:

- Psychiatric disorders
- Coronary heart disease
- High blood pressure

- Obesity
- Tooth decay
- Diabetes

Source: <https://obesity-diet.nutritionalconference.com/events-list/imbalanced-diet-effects-and-causes>
[https://www.thelancet.com/article/S0140-6736\(19\)30041-8/fulltext](https://www.thelancet.com/article/S0140-6736(19)30041-8/fulltext)


Automatic Diet Monitoring & Control - Our Vision



Automatically monitors food intake to determine if the eating is stress-eating or normal-eating.

Source: L. Rachakonda, S. P. Mohanty, and E. Kougianos, "iLog: An Intelligent Device for Automatic Food Intake Monitoring and Stress Detection in the IoMT", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. 66, No. 2, May 2020, pp. 115--124.

Diet Tracking Approaches




Food Tracking Apps

Table 1. Overview of popular food tracking approaches and their capabilities.

| App Name | Downloa ds | Reviews | Rating | Imag e | Input Method | | | Manu al | Scan ning | Spee ch | Datab ase searc h | Calori es | Nutriti on |
|----------------|---------------|---------|--------|-----------|---------------------|------------|--------------------------|------------|--------------|------------|----------------------------|--------------|---------------|
| | | | | | Food-Label in Image | | | | | | | | |
| | | | | | Auto | Man ual | Crow d Sour ced | | | | | | |
| MyFitnessPal | 50 M | 2 M | 4.6 | | | | | X | X | | | X | |
| FatSecret | 10 M | 268 k | 4.5 | | | | | X | X | | | X | X |
| My Diet Coach | 10 M | 144 k | 4.4 | | | | | X | | | | X | |
| Lose it | 10 M | 77 k | 4.4 | X | | | | X | X | | | X | |
| MyPlate | 1 M | 31 k | 4.6 | | | | | X | X | | | X | X |
| mynetdiary | 1 M | 31 k | 4.5 | | | | | X | | | | X | X |
| Macros | 500 k | 3 k | 4.5 | | | | | X | X | | | X | |
| Cron-o-meter | 100 k | 1 k | 4.2 | | | | | X | | | | | |
| Eating Habit | 100 k | 549 | 4 | X | | | X | | | | | X | |
| 21 day Fix | 100 k | 470 | 3.7 | | | | | X | | | | X | |
| Bite Snap | 50 k | 2k | 4.7 | X | | | | | | | | X | X |
| MealLogger | 50 k | 225 | 3.5 | X | | | | X | | | | X | X |
| EatRight | 10 k | 220 | 4.5 | | | | | X | | | | X | |
| Keto Meal Plan | 10 k | 19 | 2.6 | | | | | | | | X | | |
| YouAte | 10 k | | | X | | | | | | | | | |
| KudoLife | 1 k | 11 | 3.4 | | | | | | | | X | X | X |
| Calorific | 19 | | 3.2 | | | | | | | | X | | |
| Ate | | | | X | | | | ? | | | | ? | ? |
| Foodlog | | | | X | X | | | X | | | | X | |


Smart Healthcare – Diet Monitoring - iLog



iLog- Fully Automated Detection System with 98% accuracy.

Source: L. Rachakonda, S. P. Mohanty, and E. Kougianos, "iLog: An Intelligent Device for Automatic Food Intake Monitoring and Stress Detection in the IoMT", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. 66, No. 2, May 2020, pp. 115--124.

Smart Healthcare – Diet Prediction – Smart-Log



Source: P. Sundaravadivel, K. Kesavan, L. Kesavan, **S. P. Mohanty**, and E. Kougianos, "Smart-Log: A Deep-Learning based Automated Nutrition Monitoring System in the IoT", *IEEE Transactions on Consumer Electronics (TCE)*, Vol 64, Issue 3, Aug 2018, pp. 390-398.

Epileptic Seizure Has Global Impact




A seizure is an abnormal activity in the nervous system which causes its sufferers to lose consciousness and control.

- Up to 1% of the world's population suffers from epilepsy.
- Epilepsy is the fourth most common neurological disease after migraine, stroke, and Alzheimer's.
- Individuals can suffer a seizure at any time with potentially disastrous outcomes including a fatal complication called "Sudden Unexpected Death in Epilepsy" (SUDEP).


Source: <https://www.epilepsy.com/learn/about-epilepsy-basics/epilepsy-statistics>

Epileptic Seizure - Our Vision




Source: I. L. Olokodana, S. P. Mohanty and E. Kougiianos, "Krig-Detect: Exploring Alternative Kriging Methods for Real-Time Seizure Detection from EEG Signals," *2020 IEEE 6th World Forum on Internet of Things (WF-IoT)*, 2020, pp. 1-6, doi: 10.1109/WF-IoT48130.2020.9221260.

Implantable for Seizure Detection and Control



Source: <https://www.kurzweilai.net/brain-implant-gives-early-warning-of-epileptic-seizure>

Seizure Detection Methods



Consumer Electronics for Seizure Detection



Source: <https://spectrum.ieee.org/the-human-os/biomedical/diagnostics/this-seizuredetecting-smartwatch-could-save-your-life>



- Embrace2: Smart-band which uses machine learning to detect convulsive Seizures and notifies caregivers.



Source: <https://www.empatica.com/embrace2/>



- Medical grade smart watch: It detects generalized clonic-tonic Seizures and notifies physicians.

Smart Healthcare - Seizure Detection & Control



Source: M. A. Sayeed, S. P. Mohanty, E. Kougianos, and H. Zaveri, "eSeiz: An Edge-Device for Accurate Seizure Detection for Smart Healthcare", *IEEE Transactions on Consumer Electronics (TCE)*, Volume 65, Issue 3, August 2019, pp. 379--387.

Seizure Detection Approaches




Cloud Vs Edge Computing

| Cloud Vs Edge | Latency | Accuracy |
|---------------------------|---------|----------|
| Cloud-IoT based Detection | 2.5 sec | 98.65% |
| Edge-IoT based Detection | 1.4 sec | 98.65% |


Source: M. A. Sayeed, S. P. Mohanty, E. Kougiianos, and H. Zaveri, "Neuro-Detect: A Machine Learning Based Fast and Accurate Seizure Detection System in the IoMT", *IEEE Transactions on Consumer Electronics (TCE)*, Vol 65, No 3, Aug 2019, pp. 359--368.

Our Neuro-Detect : A ML Based Seizure Detection System




Source: M. A. Sayeed, S. P. Mohanty, E. Kougianos, and H. Zaveri , "Neuro-Detect: A Machine Learning Based Fast and Accurate Seizure Detection System in the IoMT", *IEEE Transactions on Consumer Electronics (TCE)*, Vol 65, Issue 3, Aug 2019, pp. 359-368.

Smart Healthcare – Brain as a Spatial Map → Kriging Methods




Source: I. L. Olokodana, S. P. Mohanty, and E. Kougianos, "Ordinary-Kriging Based Real-Time Seizure Detection in an Edge Computing Paradigm", in *Proceedings of the 38th IEEE International Conference on Consumer Electronics (ICCE)*, 2020.



Source: <http://desktop.arcgis.com/en/arcmap/10.3/tools/3d-analyst-toolbox/how-kriging-works.htm>

Spatial autocorrelation principle - things that are closer are more alike than things farther


Kriging based Seizure Detection



| Works | Extracted Features | Classification Algorithm | Sensitivity | Latency |
|------------------------------|---|--------------------------------|-------------|----------|
| Zandi, et al. 2012 [23] | Regularity, energy & combined seizure indices | Cumulative Sum thresholding | 91.00% | 9 sec. |
| Altaf,etal. 2015 [24] | Digital hysteresis | Support Vector Machine | 95.70% | 1 sec |
| Vidyaratne, et al. 2017 [25] | Fractal dimension, spatial/ temporal features | Relevance Vector Machine (RVM) | 96.00% | 1.89 sec |
| Our Proposed | Petrosian fractal dimension | Kriging Classifier | 100.0% | 0.85 s |



Source: I. L. Olokodana, S. P. Mohanty, and E. Kougianos, "Ordinary-Kriging Based Real-Time Seizure Detection in an Edge Computing Paradigm", in *Proceedings of the 38th IEEE International Conference on Consumer Electronics (ICCE)*, 2020.

Seizure Control Methods




Source: M. A. Sayeed, S. P. Mohanty, E. Kougianos, and H. Zaveri, "iDDS: An Edge-Device in IoMT for Automatic Seizure Control using On-Time Drug Delivery", in *Proceedings of the 38th IEEE International Conference on Consumer Electronics (ICCE)*, 2020.

Seizure Control Methods




Source: M. A. Sayeed, S. P. Mohanty, E. Kougianos, and H. Zaveri, "iDDS: An Edge-Device in IoMT for Automatic Seizure Control using On-Time Drug Delivery", in *Proceedings of the 38th IEEE International Conference on Consumer Electronics (ICCE)*, 2020.

Diabetes is a Global Crisis




Source: A. M. Joshi, P. Jain and S. P. Mohanty, "Everything You Wanted to Know About Continuous Glucose Monitoring," *IEEE Consumer Electronics Magazine*, doi: 10.1109/MCE.2021.3073498.


Blood Glucose Monitoring – Invasive Vs Noninvasive



Traditional – Finger Pricking



Invasive Approach – Processing Blood/Serum




Noninvasive – Wearable




Noninvasive Approach – Processing Light

Source: P. Jain, A. M. Joshi, and S. P. Mohanty, “Everything You Wanted to Know About Noninvasive Glucose Measurement and Control”, *arXiv Physics*, arXiv:2101.08996, January 2021, 51-pages.



Noninvasive Glucose-Level Monitoring




Photoplethysmogram (PPG)




Near Infrared (NIR)



Automatic Glucose Monitoring and Control - Our Vision - iGLU (Intelligent Noninvasive)




Unique Near Infrared Spectroscopy for iGLU



Source: A. M. Joshi, P. Jain, and S. P. Mohanty, A Device For Non-Invasive Blood and Serum Glucose-Level Monitoring and Control, India Patent Application Number: 202011027041, Filed on: 25 June 2020.

iGLU 1.0: Capillary Glucose



Clinically tested in an hospital.

Cost - US\$ 20
Accuracy - 100%



Source: P. Jain, A. M. Joshi, and S. P. Mohanty, "iGLU: An Intelligent Device for Accurate Non-Invasive Blood Glucose-Level Monitoring in Smart Healthcare", *IEEE Consumer Electronics Magazine (MCE)*, Vol. 9, No. 1, January 2020, pp. 35-42.

iGLU – Design Implementation

Data logging for model training, validation and testing

Processing Unit

Power Supply



Clinically tested in an hospital.

Cost - US\$ 20
Accuracy - 100%

Source: A. M. Joshi, P. Jain, and S. P. Mohanty, A Device For Non-Invasive Blood and Serum Glucose-Level Monitoring and Control, India Patent Application Number: 202011027041, Filed on: 25 June 2020.

Elderly Fall Automatic Detection is Needed to Improve Quality of Life

- Elderly Fall: Approximately a third of elderly people 65 years or older fall each year.
- Fall Caused → Over 800,000 hospital admissions, 2.8 million injuries and 27,000 deaths have occurred in the last few years.

Source: L. Rachakonda, A. Sharma, S. P. Mohanty, and E. Kougianos, "Good-Eye: A Combined Computer-Vision and Physiological-Sensor based Device for Full-Proof Prediction and Detection of Fall of Adults", in *Proceedings of the 2nd IFIP International Internet of Things (IoT) Conference (IFIP-IoT)*, 2019, pp. 273--288.


Consumer Electronics for Fall Detection

| Wearables | Drawbacks |
|---|---|
|  | <p>Apple watch: uses only accelerometers, doesn't work on low thresholds like double carpet, bathroom, hardwood floors. The user must manually select the option SOS and as a reason it fails if the person is unconscious. Users may remain on the floor with no help for large hours.</p> |
|  | <p>Philips Lifeline: Uses only accelerometers and barometric sensors for pressure changes. After the fall, the system waits for 30 sec and directly connects to help.</p> |
|  | <p>Lively Mobile by greatcall and Sense4Care Angel4: Monitors fluctuations using only accelerometers.</p> |
|  | <p>Bay Alarm Medical and Medical Guardian: Use only accelerometers. Have huge base stations limiting the usage and location access.</p> |

Issues of Existing Research


- Decisions of fall are dependent on the changes in accelerometer axes only.
- Some applications have user to give response after the fall and that can be time consuming as the user might not be conscious.
- Some applications are limited to a certain location and certain type of surroundings which add up the additional costs.
- Prediction of fall or warning the user that there might be an occurrence of fall is not provided by most of the applications.

Good-Eye: Our Multimodal Sensor System for Elderly Fall Prediction and Detection



Source: L. Rachakonda, A. Sharma, S. P. Mohanty, and E. Kougianos, "Good-Eye: A Combined Computer-Vision and Physiological-Sensor based Device for Full-Proof Prediction and Detection of Fall of Adults", in *Proceedings of the 2nd IFIP International Internet of Things (IoT) Conference (IFIP-IoT)*, 2019, pp. 273--288.


Good-Eye: Elderly Fall Detection



Good-Eye: Fall detection and prediction Accuracy - 95%.

Source: L. Rachakonda, A. Sharma, S. P. Mohanty, and E. Kougianos, "Good-Eye: A Combined Computer-Vision and Physiological-Sensor based Device for Full-Proof Prediction and Detection of Fall of Adults", in *Proceedings of the 2nd IFIP International Internet of Things (IoT) Conference (IFIP-IoT)*, 2019, pp. 273--288.

cStick: A Calm Stick for Fall Prediction, Detection and Control



Fall Monitoring + Body Vital Monitoring


Accelerometer, Gyroscope, HRV, Blood Pressure, Sugar levels, SpO2

Source: L. Rachakonda, S. P. Mohanty, and E. Kougianos, "cStick: A Calm Stick for Fall Prediction, Detection and Control in the IoMT Framework", in *Proceedings of the 4th IFIP International Internet of Things (IoT) Conference (IFIP-IoT)*, 2021.

cStick - Prototyping


- For the IoMT-Edge computing, a controller has been chosen with real time sensor data from various sensors which monitor the required parameters.

cStick: Fall detection and prediction Accuracy – 96.7%.




Source: L. Rachakonda, S. P. Mohanty, and E. Kougianos, "cStick: A Calm Stick for Fall Prediction, Detection and Control in the IoMT Framework", in *Proceedings of the 4th IFIP International Internet of Things (IoT) Conference (IFIP-IoT)*, 2021.

MyWear – A Smart Wear for Continuous Body Vital Monitoring – using ECG & EMG



Source: S. C. Sethuraman, P. Kompally, **S. P. Mohanty**, and U. Choppali, "MyWear: A Novel Smart Garment for Automatic Continuous Vital Monitoring", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. XX, No. YY, ZZ 2021, pp. Accepted on 30 May 2021.


MyWear – A Smart Wear for Continuous Body Vital Monitoring – using ECG & EMG



Source: S. C. Sethuraman, P. Kompally, **S. P. Mohanty**, and U. Choppali, "MyWear: A Novel Smart Garment for Automatic Continuous Vital Monitoring", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. XX, No. YY, ZZ 2021, pp. Accepted on 30 May 2021.

MyWear – A Smart Wear for Continuous Body Vital Monitoring – using ECG & EMG

Embedded Electrodes inside MyWear




ECG Sensor

Micro-controller

EMG Sensor

MyWear Prototype Results:


- Heartbeat Classification - Accuracy - 97%
- Myocardial Infarction (Heart Attack) - Accuracy - 98%
- Stress Level Detection - Accuracy - 97%
- Muscle Activity Detection - Accuracy - 96%
- Fall Detection - Accuracy - 98.5%



Source: S. C. Sethuraman, P. Kompally, **S. P. Mohanty**, and U. Choppali, "MyWear: A Novel Smart Garment for Automatic Continuous Vital Monitoring", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. XX, No. YY, ZZ 2021, pp. Accepted on 30 May 2021.

Smart Healthcare – Some Challenges

Smart Healthcare Architecture – Requirements



Smart Healthcare – Data Quality



Source: H. Zhu, C. K. Wu, C. H. KOO, Y. T. Tsang, Y.Liu, H. R. Chi, and K. F. Tsang, "Smart Healthcare in the Era of Internet-of-Things", *IEEE Consumer Electronics Magazine*, vol. 8, no. 5, pp. 26-30, Sep 2019.

Machine Learning Challenges



Machine Learning Issues



High Energy Requirements


High Computational Resource Requirements

Large Amount of Data Requirements

Underfitting/Overfitting Issue


Class Imbalance Issue

Fake Data Issue




Source: Mohanty ISCT Keynote 2019

Smart Healthcare - Security Challenges



Source: P. Sundaravadivel, E. Kougianos, S. P. Mohanty, and M. Ganapathiraju, "Everything You Wanted to Know about Smart Health Care", *IEEE Consumer Electronics Magazine (CEM)*, Volume 7, Issue 1, January 2018, pp. 18-28.

Health Insurance Portability and Accountability Act (HIPAA)




HIPPA Privacy Violation by Types

IoMT Device Security Issue is Scary


- Insulin pumps are vulnerable to hacking, FDA warns amid recall:
<https://www.washingtonpost.com/health/2019/06/28/insulin-pumps-are-vulnerable-hacking-fda-warns-amid-recall/>
- Software vulnerabilities in some medical devices could leave them susceptible to hackers, FDA warns:
<https://www.cnn.com/2019/10/02/health/fda-medical-devices-hackers-trnd/index.html>
- FDA Issues Recall For Medtronic mHealth Devices Over Hacking Concerns:
<https://mhealthintelligence.com/news/fda-issues-recall-for-medtronic-mhealth-devices-over-hacking-concerns>

IoMT Security – Selected Attacks



Source: V. P. Yanambaka, S. P. Mohanty, E. Kougianos, and D. Puthal, "PMsec: Physical Unclonable Function-Based Robust and Lightweight Authentication in the Internet of Medical Things", *IEEE Transactions on Consumer Electronics (TCE)*, Volume 65, Issue 3, August 2019, pp. 388--397.

IoMT Security Measures is Hard – Resource Constrained




Collectively (WMD+IMD):
Implantable and Wearable
Medical Devices (IWMDs)

Implantable and Wearable
Medical Devices (IWMDs) --
Battery Characteristics:
→ Longer life
→ Safer
→ Smaller size
→ Smaller weight

Pacemaker Battery Life - 10 years

IoMT Security Measures is Hard – Energy Constrained



Pacemaker
Battery Life
- 10 years




Neurostimulator
Battery Life
- 8 years

- Implantable Medical Devices (IMDs) have integrated battery to provide energy to all their functions → Limited Battery Life depending on functions
- Higher battery/energy usage → Lower IMD lifetime
- Battery/IMD replacement → Needs surgical risky procedures

Source: Carmen Camara, PedroPeris-Lopez, and Juan E.Tapiadora, "Security and privacy issues in implantable medical devices: A comprehensive survey", *Elsevier Journal of Biomedical Informatics*, Volume 55, June 2015, Pages 272-289.

Smart Healthcare - Safety

10 devices tied to the most reports involving death (2008-2018)



CBC NEWS

Source: Health Canada & ICIJ

Source <https://planet-report.com/canadian-advocates-call-for-all-medical-implants-to-be-registered-cbc-news/>

CENTRAL ILLUSTRATION: Cardiac-Implantable Electronic Devices: Technical and Safety Considerations

| FACTORS INFLUENCING SAFETY | SMALL POTENTIAL RISKS |
|--|---|
| MR magnet: <ul style="list-style-type: none"> • Magnet strength • Radiofrequency power • Magnet position | Heating effects: <ul style="list-style-type: none"> • Tissue injury (Mainly theoretical) Strategy to minimize risk: Lead designed to limit current induction |
| Cardiac implantable device: <ul style="list-style-type: none"> • Ferromagnetic material • Presence of reed switch • Device programming | Mechanical effects: <ul style="list-style-type: none"> • Device movement (Mainly theoretical) Strategy to minimize risk: Limitation of ferromagnetic materials |
| Leads: <ul style="list-style-type: none"> • Ferromagnetic material • Lead stability | Electromagnetic effects: <ul style="list-style-type: none"> • Altered sensing/capture • Inhibited therapies • Inappropriate therapies (No significant adverse patient outcomes) Strategy to minimize risk: Lead designed to limit current induction, replacement of reed switch with Hall sensor, temporary device reprogramming |
| Patient: <ul style="list-style-type: none"> • Patient position • Patient size | |
| Indication to scan: If the benefits outweigh the very small potential risks, MRI is acceptable | |

Miller, J.D. et al. J Am Coll Cardiol. 2016;68(14):1590-8.

Source: J. D. Miller, S. Nazarian, H. R. Halperin, "Implantable Electronic Cardiac Devices and Compatibility With Magnetic Resonance Imaging", J Am Coll Cardiol. 2016 Oct, 68 (14), pp. 1590-1598.

Smart Healthcare - Ethics




Source: <https://online.alvernia.edu/articles/ethical-issues-in-healthcare/>



Source: B. Mittelstadt, "Ethics of the health-related internet of things: a narrative review", *Ethics Inf Technol* **19**, 157–175 (2017), DOI: <https://doi.org/10.1007/s10676-017-9426-4>.

Smart Healthcare – Some Solutions


IoT/CPS Design – Multiple Objectives



Smart Cities
Vs
Smart Villages

Source: Mohanty ICCE 2019 Keynote

CPS – IoT-Edge Vs IoT-Cloud



End Security/Intelligence

- Minimal Data
- Minimal Computational Resource
- Least Accurate Data Analytics
- Very Rapid Response

Edge Security/Intelligence

- Less Data
- Less Computational Resource
- Less Accurate Data Analytics
- Rapid Response


Cloud Security/Intelligence

- Big Data
- Lots of Computational Resource
- Accurate Data Analytics
- Latency in Network
- Energy Overhead in Communications

Heavy-Duty ML is more suitable for smart cities

TinyML at End and/or Edge is key for smart villages.

Our Kriging-Bootstrapped DNN Model




HJC – Hjorth Complexity
 ENT – Signal Entropy
 PFD – Petrosian Fractal Dimension

Extracted Features

4-Layer Deep Neural Network (DNN) Model

Source: I. L. Olokodana, S. P. Mohanty, and E. Kougianos, "Kriging-Bootstrapped DNN Hierarchical Model for Real-Time Seizure Detection from EEG Signals", in *Proceedings of the 6th IEEE World Forum on Internet of Things (WF-IoT)*, 2020


Our Distributed Kriging-Bootstrapped DNN Model



Source: I. L. Olokodana, S. P. Mohanty, and E. Kougianos, "Distributed Kriging-Bootstrapped DNN Model for Fast, Accurate Seizure Detection from EEG Signals", *Proceedings of the 19th IEEE Computer Society Annual Symposium on VLSI (ISVLSI)*, 2020.


Experimental Results: Dataset A

| Models | Detection Latency |
|------------------|-------------------|
| DNN | 0.80s |
| Ordinary Kriging | 0.86s |
| Krig-DNN | 0.80s |
| Dist-Krig-DNN | 0.80s |




Source: I. L. Olokodana, S. P. Mohanty, and E. Kougiyanos, "Distributed Kriging-Bootstrapped DNN Model for Fast, Accurate Seizure Detection from EEG Signals", *Proceedings of the 19th IEEE Computer Society Annual Symposium on VLSI (ISVLSI)*, 2020.

Our Secure by Design Approach for Robust Security in Healthcare CPS



Source: V. P. Yanambaka, S. P. Mohanty, E. Kougianos, and D. Puthal, "PMsec: Physical Unclonable Function-Based Robust and Lightweight Authentication in the Internet of Medical Things", *IEEE Transactions on Consumer Electronics (TCE)*, Volume 65, Issue 3, August 2019, pp. 388--397.

IoMT Security – Our Proposed PMsec




Average Power Overhead –
~ 200 μ W or 0.2 mW

| Proposed Approach Characteristics | Value (in a FPGA / Raspberry Pi Platform) |
|---|---|
| Time to Generate the Key at Server | 800 ms |
| Time to Generate the Key at IoMT Device | 800 ms |
| Time to Authenticate the Device | 1.2 sec - 1.5 sec |


Source: V. P. Yanambaka, S. P. Mohanty, E. Kougianos, and D. Puthal, "PMsec: Physical Unclonable Function-Based Robust and Lightweight Authentication in the Internet of Medical Things", *IEEE Transactions on Consumer Electronics (TCE)*, Volume 65, Issue 3, August 2019, pp. 388--397.

Our Smart-Yoga Pillow (SaYoPillow)



Source: L. Rachakonda, A. K. Bapatla, S. P. Mohanty, and E. Kougianos, "SaYoPillow: Blockchain-Integrated Privacy-Assured IoMT Framework for Stress Management Considering Sleeping Habits", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. 67, No. 1, Feb 2021, pp. 20-29.


SaYoPillow: Blockchain Details



Source: L. Rachakonda, A. K. Bapatla, S. P. Mohanty, and E. Kougianos, "SaYoPillow: Blockchain-Integrated Privacy-Assured IoMT Framework for Stress Management Considering Sleeping Habits", *IEEE Transactions on Consumer Electronics (TCE)*, Vol. 67, No. 1, Feb 2021, pp. 20-29.


Smart Healthcare – COVID-19 Perspectives

Smart Healthcare in Pandemic – Some Roles




Source: A. M. Joshi, U. P. Shukla and S. P. Mohanty, "Smart Healthcare for Diabetes during COVID-19," *IEEE Consumer Electronics Magazine*, Vol. 10, No. 1, January 2021, pp. 66--71.

CoviLearn - Smart X-Ray Device for Automatic Initial Screening of COVID-19




Source: D. Das, S. Ghosal, and S. P. Mohanty, "CoviLearn: A Machine Learning Integrated Smart X-Ray Device in Healthcare Cyber-Physical System for Automatic Initial Screening of COVID-19", *Springer Nature Computer Science (SN-CS)*, Vol. 3, No. 2, March 2022, Article: 150, 11-pages.

Comorbidities with Pre-existing medical conditions for COVID-19




Source: A. M. Joshi, U. P. Shukla and S. P. Mohanty, "Smart Healthcare for Diabetes during COVID-19," *IEEE Consumer Electronics Magazine*, Vol. 10, No. 1, January 2021, pp. 66--71.

iGLU - Our Intelligent Non-Invasive Glucose Monitoring with Insulin Control Device




Source: A. M. Joshi, U. P. Shukla and S. P. Mohanty, "Smart Healthcare for Diabetes during COVID-19," *IEEE Consumer Electronics Magazine*, Vol. 10, No. 1, January 2021, pp. 66--71.

iLog + iGLU - Our Diet Automatic Monitoring and Control for Blood Glucose Level




Source: A. M. Joshi, U. P. Shukla and S. P. Mohanty, "Smart Healthcare for Diabetes during COVID-19," *IEEE Consumer Electronics Magazine*, Vol. 10, No. 1, January 2021, pp. 66--71.

EasyBand – Safety-Aware Mobility during Pandemic




Source: A. M. Joshi, U. P. Shukla and S. P. Mohanty, "Smart Healthcare for Diabetes during COVID-19," *IEEE Consumer Electronics Magazine*, Vol. 10, No. 1, January 2021, pp. 66--71.

EasyBand in Healthcare CPS (H-CPS)




Source: A. K. Tripathy, A. G. Mohapatra, S. P. Mohanty, E. Kougianos, A. M. Joshi and G. Das, "EasyBand: A Wearable for Safety-Aware Mobility During Pandemic Outbreak," *IEEE Consumer Electronics Magazine*, vol. 9, no. 5, pp. 57-61, 1 Sept. 2020, doi: 10.1109/MCE.2020.2992034..

CoviChain: A Blockchain based Framework for Nonrepudiable Contact Tracing in H-CPS



Source: S. L. T. Vangipuram, S. P. Mohanty, and E. Kougianos, "CoviChain: A Blockchain based Framework for Nonrepudiable Contact Tracing in Healthcare Cyber-Physical Systems during Pandemic Outbreaks", Springer Nature Computer Science (SN-CS), Vol. 2, No. 2, June 2021, Article: 346, 16-pages.

CoviChain: A Blockchain based Framework for Nonrepudiable Contact Tracing in H-CPS



Source: S. L. T. Vangipuram, S. P. Mohanty, and E. Kougianos, "CoviChain: A Blockchain based Framework for Nonrepudiable Contact Tracing in Healthcare Cyber-Physical Systems during Pandemic Outbreaks", Springer Nature Computer Science (SN-CS), Vol. 2, No. 2, June 2021, Article: 346, 16-pages.

CoviChain: A Blockchain based Framework for Nonrepudiable Contact Tracing in H-CPS

- From the front-end, Covid file is submitted to the IPFS and store it.
- Once the file is stored, the hash of the file is returned to the browser console.
- The hash generated from IPFS is stored on the blockchain, instead of the actual file.

1. User Interface




2. File Converted to Buffer



3. IPFS returning Hash




4. Confirming Metamask



Source: S. L. T. Vangipuram, S. P. Mohanty, and E. Kougianos, "CoviChain: A Blockchain based Framework for Nonrepudiable Contact Tracing in Healthcare Cyber-Physical Systems during Pandemic Outbreaks", Springer Nature Computer Science (SN-CS), Vol. 2, No. 2, June 2021, Article: 346, 16-pages.

CoviChain: A Blockchain based Framework for Nonrepudiable Contact Tracing in H-CPS


Comparing MedRec and Covichain Mining Time for MB Data



■ The times for data in MedRec are calculated assuming the mining time of the conventional Ethereum blockchain to be 13 Seconds for 1MB Data.

Source: S. L. T. Vangipuram, S. P. Mohanty, and E. Kougianos, "CoviChain: A Blockchain based Framework for Nonrepudiable Contact Tracing in Healthcare Cyber-Physical Systems during Pandemic Outbreaks", Springer Nature Computer Science (SN-CS), Vol. 2, No. 2, June 2021, Article: 346, 16-pages.


PharmaChain - A Blockchain to Ensure Counterfeit Free Pharmaceutical Supply Chain



Source: A. K. Bapatla, S. P. Mohanty, and E. Kougianos, "PharmaChain: A Blockchain to Ensure Counterfeit Free Pharmaceutical Supply Chain", *arXiv Computer Science*, arXiv:2202.02592, Feb 2022, 25-pages.

Pandemic – Trusted Food Supply Chain

How to ensure quality food through legitimate supply chain?



Source: A. M. Joshi, U. P. Shukla and S. P. Mohanty, "Smart Healthcare for Diabetes during COVID-19," *IEEE Consumer Electronics Magazine*, Vol. 10, No. 1, January 2021, pp. 66--71.

Conclusions and Future Research



Conclusions

- Healthcare has been evolving to Healthcare-Cyber-Physical-System (H-CPS) i.e. smart healthcare.
- Internet of Medical Things (IoMT) plays a key role smart healthcare.
- Smart healthcare can reduce cost of healthcare and give more personalized experience to the individual.
- AI/ML is a key component of smart healthcare.
- IoMT provides advantages but also has limitations in terms of cybersecurity and privacy.
- Smart Healthcare can be effective during stay-at-home scenario during pandemic.

Future Research

- Edge-AI for smart healthcare needs research.
- Internet-of-Everything (IoE) with Human as active part as crowdsourcing need research.
- Smart Healthcare will need robust data, device, and H-CPS security need more research.
- Cybersecurity of IWMDs needs to have very minimal energy overhead to be useful and hence needs research.
- Integration of blockchain for smart healthcare need research due to energy and computational overheads associated with it.
- Privacy-aware limited healthcare data sharing in global scale to reduce spread of pandemic outbreak.

Smart Healthcare – Reality?

- Short answer - Yes

Smart Healthcare – Hype?

- Still long way to go ...

Acknowledgement(s)

This material is based upon work supported by the National Science Foundation under Grant Nos. OAC-1924112 and HBCU-EiR-2101181. Any opinions, findings, and conclusions or recommendations expressed in this material are those of the author(s) and do not necessarily reflect the views of the National Science Foundation.