

Date of publication xxxx 00, 0000, date of current version xxxx 00, 0000.

Digital Object Identifier xxxxxxxx

Non-invasive Glucose Measurement Technologies: Recent Advancements and Future Challenges

PRATEEK JAIN¹ (Member, IEEE), **AMIT M JOSHI²** (Senior Member, IEEE), **SARAJU P MOHANTY³** (Senior Member, IEEE), **LINGA REDDY CENKERAMADDI⁴** (Senior Member, IEEE)

¹Nirma University, Ahmedabad, India (e-mail: prateek.jain@nirmauni.ac.in)

²MNIT Jaipur, India (e-mail: amjoshi.ece@mnit.ac.in)

³University of North Texas, Denton, USA (e-mail: saraju.mohanty@unt.edu)

⁴University of Agder, Kristiansand, Norway (e-mail: linga.cenkeramaddi@uia.no)

Corresponding author: Linga Reddy Cenkeramaddi (e-mail: linga.cenkeramaddi@uia.no).

ABSTRACT Diabetes is a long-term condition in which a person's body cannot break down blood sugar adequately due to a shortage of insulin. The most crucial element of health care is continuously monitoring blood glucose (BG) levels. The main concern of effective glucose monitoring equipment is based on the blood-pricking technique. However, this may not be suggested for frequent glucose measurement. The paper presents various glucose-measuring technologies. The research discusses various non-invasive glucose measurement techniques and their management using advanced medical technologies. The configuration of the precise measuring device is essential to meet the blood glucose monitoring requirements that are not invasive systems. Non-invasive glucose monitoring devices solve the issue of frequently pricking patients for blood samples for clinical tests. For the goal of continuous health monitoring, a Smart Healthcare framework would be built on the Internet-of-Medical-Things (IoMT) and a Healthcare Cyber-Physical System (H-CPS) to estimate blood glucose. The study also discusses a few consumer devices and cutting-edge methods for measuring glucose. The paper also outlines the several difficulties and open challenges with glucose prediction.

INDEX TERMS Non-invasive intelligent system, Glucose monitoring, Healthcare, Edge Computing Devices, Diabetes, Glucose Controlling Paradigm

I. INTRODUCTION

The human body uses glucose as an effective origin of energy. The Normal blood glucose (BG) range (80-150 mg/dl) is required to perform daily activities by the human [1]. However, fluctuations in the normal blood glucose range might cause complications in the body. In addition, insulin is a vital hormone the body produces for glucose balance after food consumption. Food digestion produces glucose, which provides energy for routine work in daily life. If insulin is not produced properly, excessive glucose concentrations will build up in the blood. The controlling feedback system of glucose generation and utilization in the body is examined in Fig. 1 [2]. If the development of *alpha* cells is greater than that of *beta* cells, a persistently high blood glucose level may result. This disorder prevents the body from producing enough insulin to neutralize the glucose. Diabetes mellitus is




FIGURE 1: An example of the closed-loop production and consumption of glucose [2].

a disease distinguished by increased blood glucose (beyond the normal range) in the body. The key contributing factor to

diabetes is an irregular sugar profile [3]. Non-communicable disease (NCD) and chronic disease prevalence rates have multiplied during the past several years. An estimated 20 million deaths from cardiovascular disease are recorded yearly, and high blood sugar is a critical risk factor. Additionally, those with diabetes are particularly impacted by viral pandemic epidemics [4]–[6]. Diabetes patients have increased considerably over the past several years due to obesity, poor diet, an aging population, and sedentary lifestyles. Since the number of individuals living with diabetes has more than quadrupled over the past two decades, diabetes is one of the health issues with rapid increased patients in the last decades (Refer Fig. 2) [7]. 2019 had a 9.3% global prevalence of diabetes, affecting




FIGURE 2: Global tendency of Diabetes, Assumed from [7].

over 463 million people. With a 10.2% prevalence rate, it is anticipated to increase to over 600 million people by the end of two decades and reach 700 million by the end of four decades. Prevalence is 10.8% in urban areas compared to 7.2% in rural areas. Almost half of diabetes patients are ignorant of their glycemic profile. Indeed, diabetes is on the rise globally, affecting roughly 1 in 10 individuals now. In the next ten years, more than 0.5 billion individuals are predicted to have diabetes [8]. According to research from the International Diabetes Federation (IDF), more people die from diabetes than from malaria (0.6 million), HIV/AIDS (1.5 million), and TB (1.5 million) combined. [9]. Diabetes is one of the primary chronic diseases that has a long-term effect on a person's health and quality of life. Insufficient insulin, insulin resistance, or excessive glucagon production are all regarded to be the causes of diabetes mellitus (DM), which is characterized by physiological dysfunctions and high blood glucose levels [10]. It is the most important health concern of the 21st century. Since a few years ago, Type 2 Diabetes (T2DM) has seen tremendous global expansion. Any form of diabetes can cause complications in the human body that may raise the risk of death. Hyperglycemia, also known as a high

blood glucose level, causes blood vessels to harden, which can damage kidneys, cause vision problems, and, in rare cases, even cause these organs to fail. Diabetes is linked to cardiac disease, peripheral vascular disease, and amputation of limbs. Contrarily, Type 1 Diabetes Patients (T1DM) may have low blood sugar or hypoglycemia due to high insulin administration [11]. Dizziness, over-sweating, and tiredness are the most typical signs of hypoglycemia in patients, and in the worst cases, it can result in coma and death. The diseases caused by diabetes in a picture is represented in Fig. 3




FIGURE 3: diseases caused by diabetes in humans

The signs of diabetes include frequent hunger, thirst, fatigue, changes in eyesight, persistent hunger, sudden weight loss, and the quick outflow of urine [12]. Diabetes can result in blindness, heart attacks, renal illness, lower limb amputations, and blindness if left untreated for an extended time. Diabetes would increase the risk of death and add to the cost of care at the moment of delivery and treatment. Additionally, diabetic people may become less productive at work, resulting in the inability to perform certain tasks. Diabetes can also lead to several health concerns, including eating disorders, mood disorders, anxiety disorders, and digestive troubles. It would raise the risk of mortality by about 50%. Diabetes also counts toward the cost of care for treatment. With oral medications, diabetes can be controlled in its early stages. Reducing the risk of amputation, cardiovascular disease, and high blood pressure also controls diabetes. This paper is structured with specific records, which demonstrate the measurement techniques, corresponding limitations, future challenges and vision. The statistical report of review records are presented in Fig. 4. The paper is framed with segregation of records, which are based on commercially available systems, measurement frameworks and proper descriptions of methodologies. These are represented in Fig. 5. The remainder of the article is structured as follows: Section II briefly discusses various types of diabetes along with the history of glucose-measuring technologies.




FIGURE 4: Statistical report of review records

Section III explores a summary of different blood glucose-level prediction methods. Section IV explains various available procedures for non-invasive blood glucose-level prediction. Section V explores miscellaneous data processing and training of the models for non-invasive body glucose-level monitoring. Section VI briefly explains different commercial devices for non-invasive blood glucose prediction. Section VII explores the controlling procedures for body glucose-level and various available related commercial devices. Section VIII describes how glucose level estimations and control in smart health are made feasible by the Internet of Medical Things (IoMT). The limitation and open issues of glucose-level measurements are described in Section IX. The conclusion of the work is outlined in Section X. The future work is summarized in Section XI.

II. PUBLIC HEALTH ISSUES OF DIABETES AND THE REQUIREMENT FOR GLUCOSE MONITORING

This section explains the issues of diabetes and the health problems caused by diabetes. It also describes the necessity of blood glucose level prediction.

A. LEVEL OF DIABETES

Diabetes develops due to inadequate insulin about glucose produced within the body [13]. The beta cells in the pancreas produce the body's insulin, which is either inadequate or nonexistent. Diabetes impairs the ability of the liver, muscles, and fat cells to balance glucose and insulin adequately. Three categories are used to classify diabetes: type 1 diabetes, type 2 diabetes, and gestational diabetes (Refer Fig. 6) [14]. For a person with type-1 diabetes, their immune system is weakened because the pancreas does not generate insulin within the body. As a result, they cannot produce insulin

naturally [2], [15]. When a person has type 2 diabetes, their pancreas produces insufficient insulin to stabilize their body's glycemic profile. Pregnant women typically develop gestational diabetes later in the birth process. Globally, 2 billion individuals will be overweight in 2020, with 300 million of those adults being obese. Additionally, there are at least 155 million overweight or obese children worldwide. By 2025, it is anticipated that 8% of people will have hyperglycemia, which will rise to 10% [8]. The significant rise in Type 2 Diabetes incidence at younger ages, has raised concerns for diabetic individuals, particularly in developing nations. In contrast, in nations with maximum industries, most people begin to have high blood sugar around 60 years of age. In the highly developed nations, middle-aged people between the ages of 35 and 64 are most commonly affected [7]. In India, Type-2 diabetes affected 69.2 million people in 2019. Adults with Type-1 diabetes number 2.35 million on average. In general, only around 5% of individuals are considered to be Type-1 diabetics, whereas the remaining 90% to 95% are Type-2 diabetics. Insulin is required for type-1 diabetic patients to maintain blood glucose control. Patients with type-2 diabetes can manage their blood glucose levels by adhering to an optimized diet, taking their medications, and engaging in regular physical activity.

B. THE DIABETES-RELATED HEALTH CRISIS

Diabetes is mostly brought on by an uneven level of BG-insulin in the body, where insulin cannot be produced from natural processes [15], [16]. In contrast to non-diabetes cases, the likelihood of mortality would also rise by up to 50%. Because of continuous monitoring, diabetes can be controlled with correct precautions. The main requirement is a smart healthcare system that enables real-time and precise, accurate self-measurement of blood glucose. Many health organizations worldwide have identified hyperglycemia as the primary problem [17], [18]. There are a lot of techniques for BG measurement [19]. Substantial methods have familiarised professionals and patients with the device [20]. Diabetes often develops between the ages of 18 and 80 [21], [22]. Glucose levels should be between 70 and 150 mg/dl in the normal range and 40 to 550 mg/dl in the pathological range [23]. Developing a device for measuring glucose levels without pricking is one of the challenging issues for healthcare [24]. For the past 20 years, glucose monitoring equipment has been available with features advancement [25].

C. NEED OF GLUCOSE MEASUREMENT FOR DIABETES MANAGEMENT

The glucometer is a device to measure the blood glucose level of the body. Diabetic patients can use such type of devices for frequent glucose measurement. Accordingly, they can modify their food habits and medication to keep their blood sugar levels within the desired range. There is a correlation between tight glycemic control and a lower risk of complications from diabetes. This makes it possible to customize diabetes treatment programs to improve blood sugar regulation and general




FIGURE 5: Data segregation reports from review records




FIGURE 6: Different category of diabetes with symptoms.

health. Diabetic people are the main end users of glucometers. Diabetic patients are conscious about their blood sugar levels and the variables that affect them through routine monitoring. The device would also be able to provide feedback for control mechanisms adopted by people for diabetic management.

D. A HISTORICAL OVERVIEW OF GLUCOSE MONITORING

A portable medical tool for determining blood sugar levels in the body is the glucose measurement device (also known as a glucometer) [13], [26]. Additionally, the glucose profile (different glucose measurement test) might be demonstrated using a strip-based test dipped in any substance. It is an excellent tool for measuring blood glucose in those with diabetes or hypoglycemia. Lyons and Clark from Cincinnati first proposed the biosensor in 1962 to improve glucose monitoring devices. A lot of people use glucose sensors for measurement. This glucose biosensor comprised an exterior dialysis membrane, an electrode of oxygen, a thin layered GOx, and an

inner semipermeable oxygen membrane. To create an enzyme electrode, enzymes might be drawn towards a detector [27]. Nevertheless, the fundamental drawback of first-generation glucose biosensors was the need for high hydrogen peroxide amperometric measurement operating potential for good selectivity. Second-generation glucose sensors (mediated glucose biosensors) have replaced the first-generation glucose biosensors. The suggested biosensors up to this point reflect breakthroughs in terms of device mobility and measurement precision. Nevertheless, these sensing devices weren't utilized for real-time diagnosis because of several proportions and environmental restrictions. Fig. 7 depicts the timeline of glucose measurement [28].




FIGURE 7: The Historical Evolution of Glucometer with Technology Development

E. DISCUSSED PRIOR TECHNIQUES FOR GLUCOSE MEASUREMENT

Currently, laboratory-based or self-monitoring is used for glucose monitoring. Both of these methods involve pricking of the blood, which is uncomfortable and only provides the facility to monitor blood glucose at that specific moment. Additionally, multiple-time pricking during the day is also uncomfortable for the user. Hence, numerous people are doubtful about choosing this kind of solution. Because of this process of measurement, people don't measure glucose timely. This might result in incorrect insulin doses and unplanned dietary ingredients. However, they are a reliable solution because of their excellent accuracy and strong sensitivity for measuring glucose. [29], [30].

Since a few years ago, scientists have been investigating several approaches to predict body glucose based on physical detection rather than conventional pricking approaches. The interstitial fluid (ISF) is used in this non-invasive approach instead of blood to detect glucose molecules. The difference in approaches of invasive and non-invasive measurement is represented in Fig. 8. There have been several attempts to




FIGURE 8: Invasive versus Noninvasive Glucose Measurement.

assess glucose through sweat, tears, saliva, and body skin [31]. The key difficulty is having accurate measurements with excellent sensitivity and dependability. Such a method could work well for self- and Continuous Glucose Monitoring (CGM) applications. These CGM methods provide measurements on fixed time slots during the day, which would be helpful for better controlling glucose levels and also for vital

early detection. Appropriate treatment would be possible for patients with hypoglycemia and hyperglycemia. These techniques would help the dietitian and medical representative to prescribe a proper diet plan for the individual.

F. BASIC REQUIREMENT OF CONTINUOUS GLUCOSE MONITORING

There are non-invasive, semi-invasive (or minimally invasive), and invasive methods for measuring glucose. An invasive method may result in trauma, which confirms the impossible situation of continuous monitoring. Without drawing blood, non-invasive procedures may be helpful for Continuous body Glucose monitoring. However, the most effective method for comfortably measuring blood glucose is non-invasive glucose measurement [32].

The CGM helps to analyze blood glucose levels at each mealtime properly. Fluctuation in glucose-insulin values can be seen after insulin secretion, physical activity, and prescribed medicine with diet. Regular glucose level fluctuation analysis is also helpful for the diabetologist to prescribe the required treatment. Patients with type 1 diabetes who monitor glucose values for insulin doses would benefit from CGM. Continuous monitoring makes it easier to control diets. The CGM flow diagram is presented in Fig. 9 [29], [30]. Patients can use the CGM to assess their BG level at any time. Estimating the prior three months' BG level for identifying HbA1c value would be useful.




FIGURE 9: Purposes of continuously monitoring blood sugar.

III. A GENERAL OVERVIEW OF GLUCOSE-LEVEL MEASUREMENT METHODS

An overview of the various glucose-level measuring techniques is included in this section. Much work has been done to measure BG levels using invasive, semi-invasive, and non-invasive methods. A lot of researchers have explored using non-invasive BG measurement techniques. Technically, Optical and non-optical processes stand for precise measurement. Some optical approaches used the PPG method, NIR spec-

troscopy, and Raman spectroscopy. In Fig. 10, a taxonomy of the various techniques is shown [26], [29], [30], [33].

A. INVASIVE METHODS

Commercially available electrochemical sensors are widely used in continuous blood glucose measuring systems [34]. When blood glucose molecules are detected, the value is monitored instantly using strip-based measurement [35]. The initial stage of blood glucose monitoring for various commercially available devices uses the lancets (for pricking the blood) [36]. Due to measurement, the process involves taking a blood sample from the fingertip more than three to four times a day, there is a factor of anxiety [37]. The minimally invasive biosensor for glucose monitoring was created using glucose oxidase and only needs a 1mm skin puncture to take measurements [38]. A small drop of blood drawn represents the photometric method for BG level estimation [39].

B. SEMI-INVASIVE PROCEDURES

A semi-invasive technique employing an embedded sensor has been proposed to monitor glucose tissue levels regularly. The sensor is wearable and placed on a membrane containing immobilized glucose oxidase. The development of implanted devices for glucose monitoring is presented in the paper [40]. The diabetes patient-specific semi- or minimally invasive technique employing biosensors [41]. The wearable micro system investigated for routine glucose measurement is presented as [42]. One similar method is explored for BG level estimation using a transponder chip and a micro-fabricated biosensor [43]. The Dexcom sensor's semi-invasive technique was represented for BG-level prediction using the transponder chip signal. The artificial pancreas system with glucose sensor for diabetes management is discussed in work [44]. The primary drawbacks of minimally invasive techniques are the survival of sensors for a long time and limited corresponding environmental constraints.

The portable small system checks BG levels continuously. It is a semi-invasive technique for measuring glucose. The main concept of the approach is that it employs a wearable micro-actuator that is made of shape memory alloy (SMA) to take blood samples. The implementation of PCB makes use of an enhanced version of SMA. It may be regarded as the earliest portable glucose monitoring device due to its efficacy and viability, but its size makes it uncomfortable.

C. NON-INVASIVE METHODS

All of the above-mentioned issues would be eliminated by non-invasive measurement [45], [46]. The glucose-measuring is possible without pricking through intelligent healthcare technologies [32]. Many strategies are explored for glucose estimation [47]. Compared to invasive and semi-invasive methods, non-invasive measurements are more practical for continuous glucose monitoring [45], [46]. The literature has found that glucose measurement using the optical approach is more accurate and dependable [48]. Non-invasive measurements using Raman spectroscopy, near-infrared spectroscopy,

photo spectroscopy, polarimetric, and scattering spectroscopy are among the common optical techniques [49], [50] etc. According to the researcher, the development of non-invasive measuring equipment would be considerably more practical from the user's point of view [51], [52]. Recently developed non-invasive systems are trained using serum glucose values, which provide an impressive level of precision (refer to Fig. 11). Fluctuation of the body glucose to interstitial body glucose dynamics is thought of for a continuous body glucose monitoring system's accuracy [53], [54]. For portable setup, several calibration techniques have been explored and implemented [55]. A number of deliberate attempts have been made to develop the self-monitoring system [56].

D. THE TRADE-OFFS BETWEEN INVASIVE AND NON-INVASIVE GLUCOSE MEASUREMENTS

Current glucose monitoring techniques for the growing global population of sugar patients are invasive, time-consuming, uncomfortable, and need a lot of disposable goods, which continually add to family expenses. Such restrictions are overcome by the non-invasive glucose measuring technology, an area of study that has grown substantially in recent years. However, there is a compromise between these two approaches, as seen in Fig. 12.

E. NON-INVASIVE PREDICTION FOR SERUM VS CAPILLARY GLUCOSE

When compared to the capillary glucose level, the serum glucose value is more accurate. Most traditional methods test capillary glucose rapidly using a one-touch device, but it is challenging to identify serum glucose measurements instantly. It has been seen that capillary blood glucose levels are consistently greater than serum glucose levels. An accurate blood glucose reading would aid in taking the proper management measures. Because serum glucose is more reliable for diagnostic purposes compared to capillary glucose, it is crucial to monitor it. Capillary blood glucose prediction is preferred more frequently than serum glucose estimation. Blood glucose cannot be measured continuously or frequently enough to treat diabetes. Blood sugar can be managed considerably more effectively if serum glucose can be measured regularly. Only serum blood is used for the laboratory examination of HbA1c, which offers a three-month body glucose prediction. Serum and capillary glucose levels are being monitored using optical methods for non-invasive measurements. The basis for measuring blood glucose is IR light that has been obtained after being absorbed and scattered by glucose molecules flowing through blood capillaries. Except for the post-processing computing models needed for blood glucose estimates, the methodologies for the two glucose measurement methods are relatively comparable.

F. A NON-INVASIVE METHOD FOR SALIVARY GLUCOSE LEVEL ESTIMATION

Salivary glucose measurement has been performed on adults and children [57]. There are distinct elements in saliva are




FIGURE 10: An overview of the Options for Glucose Measurement [26], [29], [30], [33].




FIGURE 11: Measurement of Serum Glucose using NIR Spectroscopy.

categorized: (1) Saliva from certain glands, and (2) total saliva. Individual glands collect the gland-specific saliva, including the parotid, sub-mandibular, sublingual, and minor salivary glands. The knowledge of salivary composition helps to diagnose diabetic patients.

Based on the patient's risk factors, family history, age, and sex, diabetes level can be confirmed after justifying the symptoms. Other glucose measurement techniques include photo-metric measurements, which only require extremely tiny sample volumes [58]. The fundamental method provides the predicted BG value using a chemical test strip. The test area's reflections are measured, and the estimated glucose level is obtained.

IV. TECHNIQUES HIGHLIGHTED FOR NON-INVASIVE BLOOD GLUCOSE-LEVEL PREDICTION

This section provides an in-depth analysis of the different body glucose monitoring methods without pricking. The methods are currently being considered for future aspects of non-invasive measurement. There have been several attempts to monitor glucose non-invasively utilizing optical methods [28], [30], [59], [60]. These methods primarily rely on a variety of spectroscopy-based approaches. The researcher believes that designing a non-invasive measuring system would be considerably more practical from the user's standpoint. Fig. 13 shows an overview of different kinds of body glucose measuring techniques without pricking. Their pictorial representation is shown in Fig. 14. A comparative and qualitative analysis of different methods without pricking is summarised in Table 1.

A. NEAR-INFRARED (NIR) SPECTROSCOPY

When infrared radiation is incident on an object, the mechanism refers to vibration spectroscopy and infrared spectroscopy (IR spectroscopy) [61], [62]. The many IR spectroscopy types are displayed in Fig. 15. Reflection, scattering, and absorption spectroscopy are the most common types of IR spectroscopy [63]. The IR absorption wave induces molecular vibrations that result in a spectrum band with a wavelength of cm^{-1} [64]. The basic data acquisition module using IR spectroscopy is represented (refer to Fig. 16).

In this instance, the item (which may be a finger or an ear lobe) is exposed to light with a near-infrared wavelength range of 700 nm to 2500 nm. [66]. According to [67], [68], it

is possible for the light to interact with blood constituents and be dispersed, absorbed, and reflected. According to the Beer-Lambert law, the received light's intensity varies with blood glucose levels [69], [70]. The receiver would aid in measuring the amount of glucose present in the blood artery, according to Nikawa et al [71]. A vernier effect-based ultra-sensitive optical sensor for measuring glucose levels is disclosed. Using the lateral offset single side-hole optical fibre (SSHOF) construction. Two integrated parallel Mach-Zehnder interferometers (MZIs) have differing free spectral ranges (FSRs). The problem is then resolved by a unique FSR regulating technique that is developed using the optical path difference (OPD) theory [72].

1) Long and Short-Wave NIRS

An effective method for measuring glucose precisely is optical detection. An optical approach based on FIR (far infrared) aids in obtaining the resonance process between O-H and C-H for the initial overtone. Long-range NIR, however, performs well in in-vitro tests. Similarly, for vitro-based glucose monitoring, a fiber-optic sensor is utilized in conjunction with laser-based mid-infrared spectroscopy. Continuous glucose measurement has been made possible using a multivariate calibration model for error analysis [73]. Compared to short-wave NIR, the FIR method has a restriction of shallow penetration. The short NIR would aid in more precise detection of the glucose molecule [74]. Fig. 17 illustrates the idea of NIR spectroscopy for glucose detection. NIR spectroscopy's unique wavelength has previously been used for accurate non-invasive glucose measurement in the past [75]. For the detection of glucose, a particular wavelength, such as 940 nm, has been investigated [76]. With NIR spectroscopy, the vibration of the C-H molecule has been seen at 920 nm [75]. In several additional studies, the validity of glucose absorption in the 1300 to 1350 nm range and the identification of glucose stretching in the NIR range [77], [78]. The existence of glucose element has been estimated at 1300 nm in related work [79].

2) Prior Approaches of NIR Spectroscopy

In the literature, a technique to calculate non-invasive blood glucose utilizing NIR spectroscopy and PPG has been proposed [80]. This technique uses a photodetector, NIR LED, and an optode pair. An analog front-end system produces a PPG signal at NIR wavelengths (935, 950, and 1070 nm). The glucose levels have been assessed using an FPGA-based Artificial Neural Network (ANN). For painless and autonomous blood extraction, a microcontroller is utilized [81]. The best solution uses Blood Glucose Measurement (BGM), which transmits and displays blood glucose data using a microcontroller. A remote device monitors the insulin pump, which is necessary for managing diabetes. Because it produces sound waves, this form of measuring technique takes advantage of changes in the intimidation of the susceptible body part [82]. The acoustic waves will respond more strongly when the glucose concentration is higher. The signal is then boosted

TABLE 1: Comparing the quality of several glucose monitoring techniques without pricking.

Approach	Benefits	Drawbacks
<i>Near Infra – Red (NIR)</i>	<ul style="list-style-type: none"> signal strength is inversely correlated with the glucose molecule. Other interacting materials, such as plastic or glass, might be used with the glucose sensing approach. 	<ul style="list-style-type: none"> Complicated machine-learning model required due to low strength signal. High scattering
<i>Mid Infra Red (MIR)</i>	<ul style="list-style-type: none"> Stronger glucose molecule absorption Low scattering 	<ul style="list-style-type: none"> The light only partially penetrates tissue. Noise in the signal because of blood composition.
<i>Far Infra Red (FIR)/ Thermal emission spectroscopy</i>	<ul style="list-style-type: none"> It is not necessary to calibrate frequently. Least susceptible to dispersion 	<ul style="list-style-type: none"> The material's temperature and thickness affect the radiation's intensity. Precision issue in glucose measurement due to strong water absorption
<i>Raman Spectroscopy</i>	<ul style="list-style-type: none"> Less susceptibility to water and temperature High specificity 	<ul style="list-style-type: none"> laser radiation not suitable for CGM Low SNR and susceptible to noise interference
<i>Photo acoustic</i>	<ul style="list-style-type: none"> Easy-to-use sensor design No tissue harm due to optical radiation 	<ul style="list-style-type: none"> Susceptible to acoustic noise, temperature, mobility, etc. Conveys some noise from some blood components that don't contain glucose.
<i>Polarimetry</i>	<ul style="list-style-type: none"> The laser intensity fluctuation won't significantly alter the predicted glucose levels. 	<ul style="list-style-type: none"> External laser source is necessary, and good eye alignment is necessary. Sensitive to changes in temperature and PH
<i>Reverse Iontophoresis</i>	<ul style="list-style-type: none"> Based on a straightforward electrode setup for enzymes Accurate glucose measurement from interstitial fluid 	<ul style="list-style-type: none"> Calibration requirement Not user-friendly solution due to the current going through the skin
<i>Fluorescence</i>	<ul style="list-style-type: none"> Due to immunity to light scattering, very sensitive to the detection of glucose molecules Due to recognizable optical characteristics, good sensitivity 	<ul style="list-style-type: none"> Very sensitive to pH and oxygen levels short life span
<i>Bio impedance spectroscopy (IMPS)</i>	<ul style="list-style-type: none"> Low-cost solution Easier for frequent measurement 	<ul style="list-style-type: none"> Sweating, movement, and temperature constraints Extended calibration duration is necessary
<i>Millimetre and Microwave sensing</i>	<ul style="list-style-type: none"> Deep penetration depth to detect glucose precisely No risk for ionization 	<ul style="list-style-type: none"> Poor selectivity Very sensitive to physiological factors including breathing, heart rate, and perspiration
<i>Optical Coherence Tomography</i>	<ul style="list-style-type: none"> High resolution and good SNR Not susceptible to the change in blood pressure or heart rate 	<ul style="list-style-type: none"> Glucose levels may fluctuate depending on skin and movements Suffers from tissue inhomogeneity
<i>Surface Plasma Resonance</i>	<ul style="list-style-type: none"> High sensitivity to detect small glucose molecules 	<ul style="list-style-type: none"> Lengthy calibrating procedure and large size Sensitive towards change in temperature, perspiration, and movement
<i>Time of flight and THz Time domain Spectroscopy</i>	<ul style="list-style-type: none"> High glucose molecule absorption High dispersion 	<ul style="list-style-type: none"> A longer measuring time Worse depth resolution
<i>Metabolic Heat Conformation</i>	<ul style="list-style-type: none"> Uses the idea of well-known physiological characteristics to predict glucose levels. 	<ul style="list-style-type: none"> Sensitive to changes in temperature and perspiration
<i>Electromagnetic sensing</i>	<ul style="list-style-type: none"> Inexpensive and readily miniaturized No risk of ionization 	<ul style="list-style-type: none"> Dielectric constant-related lack of selectivity is mostly affected by other blood components. More sensitive to even small temperature changes
<i>Ultrasound Technology</i>	<ul style="list-style-type: none"> Dependable technique that doesn't significantly injure tissue cells Extending below the skin or tissue for a long time 	<ul style="list-style-type: none"> limited precision hence multi-model approach utilized Expensive measuring technique
<i>Sonophoresis</i>	<ul style="list-style-type: none"> Favourable technique since there are no skin adverse effects. Based on a well-used enzymatic technique 	<ul style="list-style-type: none"> Constraints of Environmental factors




FIGURE 12: Glucose measurement trade-offs between invasive and non-invasive methods.




FIGURE 13: Different spectroscopy techniques for non-invasive glucose prediction.




FIGURE 14: Proximate Outlook of Different Prevalent Spectroscopy Approaches for Glucose Measure without Pricking.




FIGURE 15: Classification of vibrational spectroscopy [65].




FIGURE 16: Block diagram representation of IR spectroscopy.

to increase SNR and lower noise before being sent to the computer for additional processing. Photo-acoustic amplitude estimates feature extraction and glucose estimation. The acoustic signals are composed using double laser diodes and a PZT. The arrangement is expensive and cumbersome because the LASER is being used. A portable and precise glucose concentration measuring device is developed using the well-




FIGURE 17: Penetration depth of various Infrared Signals in Human Skin [26], [33].

known photoacoustic near-infrared spectroscopy. The investigated in-vitro instrumentation techniques have a compact form factor, making them a good option for a non-invasive, in-vivo wearable blood glucose monitoring system. But, due to some drawbacks, it could not be possible to consider commercialization [83].

3) BG Monitoring Device without Pricking (iGLU)

The prototype intelligent glucometer device (iGLU) is employed to collect data. After processing the sensor data, a sophisticated computing model is explored. Three channels have been used to implement this prototype. It stores data on the cloud and employs remote data monitoring [84]. The explored prototype uses dual short wavelengths and a NIR Spectroscopy [85]. It collects the sensor data in three different ways. For optical detection, each channel includes a separate emitter and detector. An ADC with a suitable sampling rate has been used to handle the data for estimating BG values. Regression analysis examines the optimized model and calibrates and validates the data. Patients and physicians may utilize and keep an eye on the data that is kept in the cloud. Based on the values of the saved data, treatment may be offered. This is a low-cost solution with greater than 90% accuracy.

4) Why are NIR Methods preferred Over Other Noninvasive Methods?

Numerous non-invasive techniques have been used to monitor glucose, including IMPS, NIR light, and PPG signal processing. However, other methods, except optical detection, haven't been able to deliver accurate measurements. One

better option is PPG. Although the PPG signal changes depending on blood concentration [86], [87]. Accurate blood glucose predictions might not be helpful in exceptional cases. It means that the device should measure the precise BG of all people. Each person has distinctive characteristics for their sweat and saliva. As a result, it could not be an accurate method for measuring glucose. Other spectroscopy has also been used to measure glucose levels. Nevertheless, they cannot offer reliable, affordable, or portable predictions of body glucose. The optical detection of glucose utilizing long NIR waves has shallow penetration, making it unable to detect glucose molecules under the skin [74]. Small NIR waves have been considered a viable remedy for real-time glucose measurement [76], [88].

B. MID INFRA-RED (MIR) SPECTROSCOPY

Mid-infrared (MIR) spectroscopy provides high resonance of glucose molecules comparatively [89]. Due to the skin's propensity for greater water absorption, skin penetration depth is quite low. This method aids in the determination of ISF glucose in vivo. There have been various attempts to detect glucose accurately using samples from the palm and saliva.

C. PRIOR PPG APPROACH FOR BLOOD GLUCOSE MEASUREMENT

Through the PPG signal, the change in blood volume caused by the tissue's absorption of light has been discovered [87]. With the use of a light detector and pressure pulse, the change in blood volume has been detected [86]. It is possible that the glucose molecule is not the cause of the change in blood volume since it would be equivalent to a change in light intensity. This might lead to incorrect glucose readings. The difference between NIR and PPG is seen in Fig. 18. The NIR spectroscopy principle is the primary method of glucose measurement. The intelligent glucose measuring device iGLU is explored for accurate BG value prediction. There have been several studies for PPG signal-based glucose detection [90]. The patient's body data has been stored to assess the existence of glucose molecules using PPG. Different machine-learning algorithms have since been applied to estimate the value of body glucose [91]. Using ARMA models, the various parameters from a total of 70 healthy and diabetic participants have been considered for the prediction [92]. There have also been several other methods for estimating glucose utilizing PPG signals [93]–[95].

Photo-plethysmography (PPG), an optically based technology, is employed in cutting-edge medical treatment. It is a method of measuring glucose without pricking. The PPG signal is recorded using a sensor comparable to a pulse oximeter [87]. The detector, which will work in the NIR area, is constructed using a photo transmitter and receiver. A PPG signal may be measured at wavelength 920 nm by monitoring variations in light absorption. With each heartbeat, the veins in the finger expand and contract. Blood flow has been observed based on this mechanism, and glucose level




FIGURE 18: Comparison of PPG and NIR for Glucose Prediction without Pricking [26], [33].

estimation is explored. There is a way to measure blood sugar using an oximeter device and transmit the PPG signal for glucose monitoring system [90]. The signal is first acquired as a photo-current, which is then converted into quantifiable voltage values for filtering. Lab View calculates the body glucose status after processing the filtered signal. A prototype of a system that uses machine learning methods and a PPG system to assess body glucose statuses has been created without pricking [86]. In this prototype, a sensor, a training detector, and a signal conditioning part are utilized to formulate the PPG data. The functional relationship between the PPG waveform's shape and the BP and glucose levels may be established. PPG will adjust the light intensity depending on variations in blood volume. The analysis of PPG signals does not demonstrate glucose molecule detection. Consequently, the system's precision is restricted [26], [33]. Fig. 18 demonstrates the distinctions.

D. IMPS APPROACH

Dielectric spectroscopy is known as impedance spectroscopy (IMPS) [96]. Fig. 19 depicts the stages involved in impedance spectroscopy (IMPS). This method determines the skin's dielectric characteristics [97]. The skin is exposed to the current [98]. The impedance range is acquired as a result of the directed tiny current at various wavelengths [99]. The frequency range is 100 Hz to 100 MHz [100], [101]. Changes in sodium ions and potassium ions concentration will be reflected in changes in glucose concentration [102]. Thus, there will be a shift in the cell membrane potential differential [103]. As a result, the dielectric value will alter and forecast the human body's glucose level [104].

The detection of salivary glucose using a sensor of detecting enzymes in a cell has been investigated [105]. To detect glucose in human saliva, polypyrrole (PPy) supported with copper (Cu) nanoparticles on an alkali anodized steel (AS) electrode is available in [106]. Because each person's




FIGURE 19: Representation of Steps of IMPS.

sweat and saliva have different qualities, a high accuracy level cannot be achieved using these procedures. This method cannot be used to monitor glucose in smart healthcare.

E. RAMAN SPECTROSCOPY

A glucose molecule's polarisation will alter due to light's interaction with it [107]. Through the use of LASER light, this method makes it feasible for the solution's molecules to rotate and oscillate [108]. The molecule's vibration influences the scattered light emission [109]. This idea allows for predicting blood glucose levels as [110]. When compared to infrared spectroscopy, this method offers more precision [111]. Numerous studies using Raman spectroscopy have been conducted to detect glucose precisely. "In-vivo" measurement has also been used for the validation. Fig. 20 outlines the fundamentals of Raman spectroscopy, while Fig. 21 explains how to use it to measure glucose without using a needle.




FIGURE 20: Building blocks of Raman spectroscopy.

F. TIME OF FLIGHT AND THZ DOMAIN

Time of Flight (TOF) measurements are used to estimate blood glucose levels during in vitro measurement [112]. A brief laser pulse is introduced into the sample to measure photon migration. While leaving the sample, this photon will encounter scattering and absorption phenomena. The optical




FIGURE 21: Noninvasive glucose measurement using Raman spectroscopy.

analysis of the photons would be beneficial for accurate glucose measurement.

G. PHOTO ACOUSTIC SPECTROSCOPY

The photo-acoustic technique is the acoustic effect for assembling the acoustic strain wave from an object (direct Fig. 22) [113]. The estimate of blood glucose detection in this spectroscopic method is provided by the absorption of modulated light input [114]. According to its optical circumstances, an item absorbs high-intensity optical light [115]. This procedure enabled excitations of certain molecules based on their resonance frequencies [116]. The absorbed light is seen as heat, causing a rise in the sample's local temperature and thermal expansion [117]. The increase in volume creates acoustic pressure [118]. Through particular keen wavelengths that are vibrant for the vibration of glucose molecules, the resulting photoacoustic wave may be utilized to estimate the glucose concentration [119]. The glucose molecule modifies its properties at a certain resonance frequency. The acoustic waveform has changed [120]. Previously, optical light with a wavelength of 905 nm was utilized for excitation [121], [122].




FIGURE 22: Photo acoustic spectroscopy.

H. CAPACITANCE SPECTROSCOPY

The inductor stray capacitance in the capacitance spectroscopy technique changes with body capacitance (Fig. 23)

[123]. To gauge body glucose levels, one uses the body capacitance [124]. A flexible inductor-based sensor uses the coupling capacitance concept to detect body glucose. The inductive sensor and skin of the body do not come into contact in this method due to the current [125]. Depending on body glucose, the inductive sensor's stray capacitance will change. With this method, the impact of fat and muscle on blood glucose will be minimal [126].




FIGURE 23: The typical steps of capacitance spectroscopy.

I. SURFACE PLASMON RESONANCE (SPR)

The Surface Plasmon Resonance (SPR) uses an electron oscillation technique at the metal and dielectric contact to detect glucose [127]. It primarily picks up changes in refractive index before and after analyte contact. Due to its mobility, the optical fibre-based SPR is employed for glucose PoC monitoring.

J. RF & MICROWAVE SENSING APPROACH

The fluctuation in the s-parameters response in the RF method directs the difference in body glucose [128], [129]. Fig. 24 illustrates specific stages of this approach. The output is determined via the antenna or resonator [130], [131]. They observe the transitions in dielectric constant value via the dispatch [132]. Through the antenna or resonator, the change in resonance frequency spectrum may be used to detect changes in the dielectric constant [133], [134]. Blood dielectric constant changes according to the distinct blood glucose levels. The human finger is a suitable measuring object, but the measurement accuracy is affected because of method limitations and constraints. These include the consistency of the skin, fingerprints, pressure exerted by the object during measuring, and object placement on the sensor [135].

Reduced flexible antenna size prevents installation issues and skin-crumbling effects on the signal acquisition region. Furthermore, an effective antenna sensor design is necessary for continuous diabetes monitoring. An adjustable antenna




FIGURE 24: Glucose measurement using RF sensing technique.

detector positioned within the thumb spica splint glove has created a continuous glucose monitoring device [136].

K. OCULAR SPECTROSCOPY

Through the tears, glucose concentration is assessed using the Ocular Spectroscopy method. To forecast the body's glucose content, a certain lens is used [137]. The lens has a hydrogel wafer put on it. This wafer's $7\ \mu\text{m}$ thickness was achieved using boronic acid preparation. After the wafer has been placed on the lens, optical rays are introduced to the lens. The wavelength of the reflected light will then alter. A difference in light range will be interpreted as a difference in the amount of glucose present in tears.

L. IONTOPHORESIS

A tiny electric current diffusely flows through the skin during the iontophoresis or ionization procedure. For the same procedure, three electrodes are employed [138]. The various behavioral electrodes get a tiny voltage through the electrodes. Glucose is transported to the cathode during this process. The working electrode may perform the biosensing function by producing current during involved voltage via electrodes. This biosensor measures the body's glucose passively. The wrist is commonly used for measuring [139].

M. OPTICAL COHERENCE TOMOGRAPHY

The reflectance spectroscopy principle is the foundation for the optical coherence tomography method. In this method, the sample (which is positioned in an interferometer) is stimulated using low coherent light. A moving mirror is set up in the reference arc of an interferometer. Conversely, a photodetector is positioned to pick up the interferometric signal. The light in this signal has been reflected and backscattered. We were able to obtain excellent 2-D photos thanks to this approach. In interstitial fluids, the glucose concentration rises as the refractive index rises. The scattering coefficient

changes as the refractive index changes [107]. So, indirectly, the scattering coefficient and glucose levels are related.

N. POLARIMETRY

With greater precision, the polarimetry technique is frequently utilized in clinical laboratories. For glucose monitoring, the optical linear polarization-based method is employed [140]. This method is often based on the trajectory of a vector caused by the consistency, temperature, and BG content. Polarised light is passed through the path, retaining glucose molecules due to glucose prediction. Increased dispersal via the skin makes it feasible for the beam to depolarize. A polarimetric trial is performed using the eye to overcome this limitation. Due to the blood's temperature and pH level rotation, this procedure is completely affected [141]. The measurement process is represented in Fig. 25.




FIGURE 25: Non-invasive glucose measurement using Polarimetry.

O. THERMAL EMISSION SPECTROSCOPY

The thermal emission technique is based on the body's infrared wave, which is created naturally. The body's glucose level will influence the IR waves that are released. The tympanic membrane of the human body's typical mid-IR emission is modified by tissue emitting. This method's selectivity is identical to that of absorption spectroscopy. This method allows for determining glucose through the boneless part of the body. A similar approach is used for measuring glucose in the paper [142]. Measuring heat emission from the tympanic membrane might offer an accurate and acceptable practical solution in the clinical setting.

P. ULTRASOUND

Like the reverse iontophoresis method, the ultrasound approach relies on lower-frequency segments to remove molecules from skin [143]. Additionally, it has greater skin

permeability than reverse iontophoresis and is similar to sonophoresis. For glucose to be taken externally via the skin, an ultrasonic exposure lasting a few minutes to several minutes is required. Few attempts have been made to utilize this kind of technology, and no commercial devices have been developed.

Q. METABOLIC HEAT CONFORMATION (MHC)

MHC procedure aids in measuring the value of glucose combined with metabolic heat and oxygen levels, as well as numerous physiological parameters consideration [144]. The computing prototype for metabolic energy preservation is altered by considering various physiological characteristics, including blood flow volume, heat metabolic rate, oxygen saturation in the hemoglobin, and pulse rate. In tests on people, this technique demonstrated high repeatability and passable accuracy.

R. FLUORESCENCE

The Fluorescence approach relies on UV radiation's ability to excite blood vessels at certain frequency ranges [145]. The next step is the detection of fluorescence at the designated wavelength. Using fluorescence through tears, glucose may be detected. Visible light diffraction was used for this purpose. A UV LASER excites the glucose solution medium at 380 nm. Estimates of fluorescence, which is directly related to glucose content, were made. The signal is unaffected by changes in light intensity across the surroundings while using this approach.

S. KROMOSCOPY

The Kromoscopy method uses four distinct wavelengths and diverse spectroscopic responses to NIR light [146]. It uses a multi-channel method with overlapping band-pass series filters to identify the glucose molecule. In this procedure, the sample is exposed to IR radiation, which is then distributed among four detectors equipped with band-pass filters. Each detector will pick up the light from the tissue's identical structural features. The glucose concentration was then determined using a complicated vector analysis.

T. ELECTROMAGNETIC SENSING

Changes in blood glucose content allow the Electromagnetic Sensing technique to detect differences in blood sample conductivity [147]. Every time the blood glucose concentration changes, an electromagnetic sensor will monitor the alternation of the electric field. This approach makes use of the blood samples' dielectric parameters. The electromagnetic sensing frequency band lies between 2.4 and 2.9 MHz. The glucose molecule is most sensitive at a certain frequency that is ideal for the medium's temperature. An enduring, positively discreet, and semi-invasive implant-type electromagnetic sensor is investigated for continued body glucose monitoring that can follow differences in the glucose level [148].

U. BIOIMPEDANCE SPECTROSCOPY AND DIELECTRIC SPECTROSCOPY

The conductivity and permittivity of the membrane of red blood cells can be used to monitor blood glucose fluctuation [149]. Bio-impedance spectrum measurements are made in the 0.1 to 100 MHz frequency range. It aids in determining the resistance to electric current flowing through human biological tissue. The alteration of plasma glucose would enable the alteration of potassium and sodium to have the alteration of conductivity of the red blood cell membrane. The multi-sensor technique is typically used with this spectroscopy for accurate glucose measurement to monitor sweat, moisture, movement, and temperature. A bio-impedance transducer is proposed for noninvasive monitoring of insulin bio-availability following subcutaneous injection. The local impedance shift caused by the medication evaporating from the injection volume is used to measure insulin bio-availability indirectly. Employing a second-order polynomial function, an accuracy of $9\mu\text{l}$ was accomplished in a healthcare application environment. The anticipation was $4.2\mu\text{l}$, which is much less than the usual amount of one insulin unit ($10\mu\text{l}$) [150].

V. REVERSE IONOSPHERESIS METHOD

A tiny DC current is carried from the anode to the cathode on the skin's surface to create interstitial fluid (ISF). Iontophoresis is used for ionized molecule penetration at the skin's surface by such a low current [151]. Electroosmotic flow over the skin carries the electric potential from the anode and cathode. This would make it possible to extract the molecules via the skin as the glucose molecules are transported toward the cathode. Through the oxidation process, the enzyme technique aids in detecting the concentration of glucose molecules. The technique is well-liked and has a decent chance of measuring glucose levels accurately.

W. SONOPHORESIS

The Sonophoresis method is established on the cutaneous permittivity of the interstitial fluid (ISF) [152]. Additionally, the enzyme technique is used to test glucose. Glucose molecules have been placed on the skin's surface using a low-frequency ultrasonic pulse. The ISF's cutaneous permittivity is raised to enable glucose at the skin. The stratum corneum experiences contraction and expansion, which opens the ISF channel. This approach has been used in certain attempts to detect glucose, although it has been noted that it may be more useful for medication administration than for glucose monitoring.

X. OCCLUSION SPECTROSCOPY

The approaches based on occlusion spectroscopy rely on the idea that light scattering is inversely proportional to glucose content [153]. By exerting pressure with a pneumatic cuff, the flow is stopped for a brief period of time. The pulse brought on by the pressure excursion would cause the blood volume to alter. The light is passed through the sample, and

TABLE 2: Techniques Comparison with prior Works [26], [33].

Prior Work	Sensing Technology	Response	Distinctive wavelength	Estimating range	Linearity (%)
Kundu, et al. [160]	Opto-chemical	-	-	70-400 mg/dl	80
Sun, et al. [161]	NIR	-	NIR	60-450 mg/dl	95
Wei, et al. [162]	PPG	-	-	-	86
Singh, et al. [163]	Optical	-	-	32-516 mg/dl	80
Mayuko, et al. [164]	NIR Images	-	700-1650nm	-	-
Lee, et al. [165]	PPG	NIR	550nm-1550nm	-	75
Song, et al. [79]	Impedance and Reflectance	NIR	850-1300 nm	80-180 mg/dl	-
Pai, et al. [166]	Photoacoustic	NIR	905 nm	upto 500 mg/dl	-
Dai, et al. [167]	Bioimpedance	-	-	-	-
Beach, et al. [168]	Biosensing	-	-	-	-
Ali, et al. [88]	Transmittance and Refraction	NIR	650 nm	upto 450 mg/dl	-
Haxha, et al. [76]	Transmission	NIR	940 nm	70-120 mg/dl	96
Jain, et al. [169]	Absorption and Reflectance	NIR	940 nm	80-350 mg/dl	90
Jain, et al. (iGLU 1.0) [29], [33]	Absorption and Reflectance	NIR	940 and 1300 nm	80-420 mg/dl	95
Jain et al. (iGLU 2.0) [26], [30]	Absorption and Reflectance	NIR	940 and 1300 nm	80-420 mg/dl	97

the intensity fluctuation in the received light determines the glucose concentration. The brief delay in blood flow increases the received signal's SNR value. As a result, there would be an improvement in sensitivity for glucose detection and strong resilience for precise glucose measurement.

Y. SKIN SUCTION BLUSTER (SSB) TECHNIQUE

The Skin Suction Bluster method takes advantage of the idea of blister formation using vacuum suction over a small region of skin [154]. On fluid that has been taken from the blister, glucose is measured. Although it contains fewer glucose molecules than plasma, it is still enough for measuring glucose. This procedure is painless, well-tolerated, and has a low risk of infection. The HbA1c value, which indicates the three-month average glucose value, is always helpful to diabetic patients.

Z. MULTIMODAL APPROACH BASED MEASUREMENT

For increased-level reproducibility of blood glucose measurement without pricking, 2 concurrent spectroscopy merging IMPS and mNIR spectroscopy is investigated [155]. To overcome each technique's limitations, these two approaches are blended [156]. To estimate the glucose level, an IMPS-based design predicts the skin dielectric constant using a resonant circuit [157]. The mNIR spectroscopy technology is employed to increase NIR spectroscopy accuracy. Three wavelengths—850, 950, and 1300 nm are employed in this approach [158]. IMPS and mNIR are merged by an ANN using a processor to provide precise and accurate measurements [79]. As a result, multimodal techniques have been researched in the literature for accurate glucose measurement (Fig. 26). [83], [159].

V. CALIBRATING AND POST-PROCESSING METHODS FOR NON-INVASIVE BODY GLUCOSE LEVEL MONITORING

The post-processing and calibration methods are presented in this Section. The methods are used in various frameworks or systems for non-invasive BG level measurement.




FIGURE 26: Multimodal Design established non-invasive body glucose estimation.

A. AFTER SENSING TRAINING METHODS

An increased degree of precision and noise deduction from the acquired signal has been achieved by applying several calibration techniques. The model for errorless continuous monitoring is designed using these post-processing approaches [170], [171].

1) Performance Optimization using Signal Acquisition

The variance of random noise has been reduced using an adaptation of the coherent averaging approach [172]. By averaging N different individual samples from continuous frames, the influence of noise is reduced [173]. To boost SNR, frames with the highest count have been selected for

TABLE 3: Comparison of prior Works analytically with parameters [26], [33].

Works	R value	MARD (%)	AvgE (%)	MAD (mg/dl)	RMSE (mg/dl)	Samples (100%)	Used model	Measurement sample	Device cost
Lee, et al. [165]	-	-	-	-	9.46	A&B	Human	Blood	Moderate
Kundu, et al. [160]	0.78	-	-	-	-	-	Animal	Blood	Costly
Mayuko, et al. [160]	-	-	-	-	15.24	-	Human	Image	-
Sun, et al. [161]	-	7.31	-	-	21.06	A&B	Human	Blood	-
Wei, et al. [162]	0.87	11.33	9.8	-	-	A&B	Human	Blood	-
Singh, et al. [163]	0.80	-	-	-	-	A&B	Human	Saliva	Cheaper
Song, et al. [79]	-	8.3	19	-	-	A&B	Human	Blood	Cheaper
Pai, et al. [166]	-	7.01	-	5.23	7.64	A&B	in-vitro	Blood	Costly
Dai, et al. [167]	-	5.99	5.58	-	-	-	in-vivo	Blood	Cheaper
Beach, et al. [168]	-	-	7.33	-	-	-	in-vitro	Solution	-
Ali, et al. [88]	-	8.0	-	-	-	A&B	Human	Blood	Cheaper
Haxha, et al. [76]	0.96	-	-	-	33.49	A&B	Human	Blood	Cheaper
Jain, et al. [169]	0.90	5.20	5.14	5.82	7.5	A&B	Human	Blood	Cheaper
Jain, et al. (iGLU 1.0) [33]	0.95	6.65	7.30	12.67	21.95	A&B	Human	Blood	Cheaper
Joshi et al. (iGLU 2.0) [30]	0.97	4.86	4.88	9.42	13.57	Zone A	Human	Serum	Cheaper
Joshi et al. (iGLU 3.0) [159]	0.98	5.34	5.64	8.3	12.7	Zone A	Human	Capillary & Serum	Cheaper

averaging [174]. This suggested coherent averaging has been employed extensively through MATLAB and the coherent averaged signal collected. It has been suggested to calibrate measured data using the Golay code. Golay code implementation has resulted in the filtering or elimination of odd-measured data [175]–[177].

2) Computation Models for glucose Measurement

Several researchers have suggested the regularized least square regression model for measurement [178]. Photoacoustic signals are used to determine the estimated value. The calibration for estimating glucose concentration is done using these photoacoustic signals [179]. A multivariable linear regression model can be used to achieve this [180]. It is suggested to use a post-processing SVM approach to achieve high levels of accuracy [181]. A glucose monitoring system's SVM is a superior alternative for accurate measurement [182]. The use of ANN for data merging has also been suggested [183]. The suggested neural network model combines the measured data from several approaches [184]. The DSP processor has been involved in implementing artificial neural networks [79], [185]. To aggregate and calibrate data for the final estimated glucose concentration, the proposed computing model has been employed [186]. The implanted device, which contains low-price and power-optimized system networking and cutting-edge computing for BG monitoring, has to be implemented as a portable commercial device. An explored optimized neural network is used to illustrate a novel deep learning model [187]. An optimized non-invasive system design, a NIRS technology with specified wavelengths, and physiological characteristics are used to forecast the precise glucose value. Using the DNN model, the proposed system demonstrated an accurate model with MARD and AvgE of 12.50% and 12.10%, respectively. R2 coefficient of determination was found to be 0.97 [188].

B. MODEL VALIDATION METRICS

The calibration procedure is employed for measurements to provide accurate blood glucose values [189]. The measured glucose concentrations are compared to the obtained glucose concentration values [190]. The CEG analysis has been regarded as the most accurate measuring method for evaluating the performance of any equipment [191]. The process flow is explored in Fig. 27.




FIGURE 27: Metrics Model for System Precision measurement.

C. ANALYSES OF REAL-TIME PRECISION USING THE CEG METHOD

The clinical accuracy for biomedical applications has been examined using the Clarke Error Grid as a benchmark tool. It corrects for the physiological time delays involved in measuring body glucose and offers a point and rate accuracy prediction. The design and improvement of precise biomedical equipment will be made substantially simpler by utilizing Clarke error grid modeling. C.G. Clark created this approach in 1970 to assess the reliability of clinical trials and aid in comparing predicted blood glucose levels to blood glucose values obtained using more traditional methods. Diabetes treatment began to take notice of the Error Grid Analysis in 1987. Zones A, B, C, D, and E divide the grid's five separate zones. If the data fall into either zone A or B then the device represents clinical accuracy at the desired level. It means that the particular device, which is on trial, that




FIGURE 28: Clarke error grid analysis.

can be considered for clinical use. The Beckman analyzer has accurately or adequately forecasted the glucose findings. Zone C values could cause unnecessary adjustments, resulting in a bad result. If the readings are in zone D, it denotes a dangerous failure to measure glucose through the technology. Zone E displays the "incorrect treatment" [192]. The different zones are represented using CEG analysis in Fig. 28. CEG analysis is visualized for the representation of accuracy. This represents the correlation between the actual glucose value and the predicted glucose value for every instance.

VI. COMMERCIAL DEVICES RELATED TO GLUCOSE MONITORING

Numerous non-invasive glucometers needed for correct treatment are available on the market (e.g., the Freestyle Libre sensor and SugarBEAT from Nemauro Medical). Similar to skin patches, they would include a daily disposal function and adhesive to enable continuous body glucose monitoring. Since most commercial goods cannot monitor glucose accurately,

they are often used to treat diabetes. Products like DiaMon Tech, Glucowise, Glucotrack, Glutarac, and CNOGA medical equipment are available. Although Glutarac is a clever medical equipment, it accurately measures blood sugar. While the accuracy is still unacceptable, the cost is greater. The CGM has been implemented using the non-invasive Omelon B-2 stripless device. Glucosense, a fluorescence technique-based product, has been developed to monitor glucose levels continuously. Texas University created a flexible textile-based biosensor for measuring glucose levels. Every accessible technology has poor accuracy and is expensive.












A. COMPARING WEARABLE AND NON-WEARABLE GLUCOSE MONITORS

In the literature, both non-wearable and wearable systems have been tried for glucose monitoring. Most non-wearable methods rely on various spectroscopies, including photoacoustic, Raman, etc. The semi-invasive implanted devices are mostly biosensors in nature. Wearable devices include sweat patches, glucowatches, and smart contact lenses. LifePlus has created a wearable, non-invasive CGM device that is being considered for public purposes. Many non-invasive types of equipment are portable and useful for routine monitoring [13]. Continuous glucose monitoring is more adequate if blood glucose levels can be measured daily. Therefore, wearable technology has more cutting-edge solutions than non-wearable technology.

B. CONSUMER PRODUCTS FOR NON-INVASIVE GLUCOSE MEASUREMENT

There are several goods available such as GlucoTrack[®], glucometer from Labiotech [193], and similar methods are likewise expensive and have accuracy problems. The glucowise is yet another non-invasive device for continuous glucose measurement from Medical Training Initiative (MTI)[™]. Additionally, 2M Engineering has created the non-invasive solution based on Raman scattering spectroscopy [194]. These devices are uncommon because of their high price and high accuracy. In addition, the high degree of glucose-measuring accuracy, Glucotrack[™] has been devised by integrity applications Ltd. [195]. Three non-invasive glucose monitoring techniques were employed in succession: electromagnetic, thermal emission, and ultrasonic spectroscopy. Combining three methods provided precision and accuracy comparatively [196]. The results of comparing many consumer goods for noninvasive glucose measurement are summarized in Table 4. Different glucose monitoring-based commercial devices were not only explored to represent the corresponding methodologies and approaches but the devices are represented with their observations and limitations as well. The technical details of methodologies used have already been explained, which were used to develop the commercial devices of glucose monitoring.

TABLE 4: A Comparative perspective of a specified commercial product for noninvasive glucose monitoring.

Organization	device	Technique	Sample	Overview	Image
Cygnus Inc. (USA)	Glucowatch G2 Biographer	Reverse iontophoresis	Wrist skin	It would be worn as a watch with a disposable component, an auto sensor that would be connected to the back of the wearer's shirt and make skin-to-skin contact to check their blood sugar continuously. This was approved by the FDA on August 26, 2002.	
CNOGA (Israel)	Combo glucometer	Tissue photography analysis	Finger	This device can quickly analyze a variety of bio factors via tissue photography analysis from fingertip capillaries. This is CE approved but still pending for FDA approval.	
Pendragon Medical (Switzerland)	Pendra	Impedance Spectroscopy	Wrist Skin	The erythrocyte membrane's ability to transport sodium aids in measuring glucose; changes in transmembrane sodium fluxes are caused by impedance fields, which are recognized by the device to provide the final glucose result. This is not clinically validated.	
OrSense Ltd. (Israel)	OrSense NBM-200G	Occlusion Spectroscopy	Fingertip skin	It is based on the optical principle of a finger coupled to a sensor probe with a ring shape. The probe comprises a light source and detector for the red/near-infrared (RNIR) spectral band. It has pneumatic cuffs that provide systolic pressure and an optical signal for measuring blood sugar. This is not clinically validated.	
C8 Medisensors (USA)	C8 Medisensor Glucose detector	Raman Spectroscopy	Fingertip skin	This method is based on the detection of dispersed light after a monochromatic light source passes through the skin. The colours produced by Raman spectra aid in pinpointing the chemical composition of the glucose molecule. FDA approved it in 1999.	
Integrity Applications (Israel)	Glucotrack	Combining of Electromagnetic, ultrasonic and Thermal	Ear lobe tissue	Three distinct methods are combined in this device to boost precision and accuracy. It is approved by CE and KFSA. It is in the early stage of FDA approval.	
Tech4Life Enterprises (USA)	Non invasive glucometer	Infrared Spectroscopy	Finger	Patients with hyperglycemia or those who are pre-diabetic are benefited from routine monitoring of accurate blood glucose measurements every 30 seconds. This is not clinically validated.	
MediWise Ltd. (United Kingdom)	GlucoWise	Radio Wave Spectroscopy	Forefinger skin/Earlobe	This wireless non-invasive device measures glucose levels immediately. For the purpose of detecting blood glucose, it is based on electromagnetic waves with certain frequencies. It makes use of a thin-film metamaterial layer to improve penetration for accurate glucose measurement. FDA approval is still awaited	
Nemauro Medical (United Kingdom)	SugarBeat	Reverse iontophoresis	Arm, Leg and abdomen	This painless, continuously accurate blood glucose monitor has been demonstrated to be effective. SugarBEAT® offers needle-free, real-time glucose measurement. Typically, calibration requires a single finger-prick test. When a new patch is needed, a single finger prick is performed. CE-mark approved. FDA approval is still pending.	
Abbott Ltd. (USA)	Free Style Libre	Glucose oxidase method	Fore-arm skin	To measure glucose levels through interstitial fluid, enzyme glucose-sensing technology is employed. Through the use of a sensor, the glucose oxidase technique allows for the measurement of both glucose and an electrical current proportional to the concentration of glucose. It is recently FDA approved.	
VivaLyf (India)	EzLyf	Near Infra Red	Finger	A beam of light would interact to glucose molecules of the finger for absorption measurement then glucose is measured through the spectrum between the molecules and the wave. FDA approval is still pending	

VII. METHODS FOR GLUCOSE LEVEL CONTROL AND CONSUMER GOODS

Different models for diet regulation have been created utilizing different glucose-insulin balance factors. To predict the amount of glucose consumed by diabetes patients, the metrics primarily comprise NHGB, RTG, peripheral glucose utilization, and glucose absorption rate. These are helpful variables for calculating blood glucose levels using the right insulin dosage and a predetermined eating schedule. The blood glucose-insulin control model was created to balance the body's glucose and insulin levels in diabetes patients using the right medications [197].

A. DIABETES CONTROLS TECHNOLOGIES

The decision models for deliverable insulin administration are provided to demonstrate the blood regulation factors. A model of insulin secretion and glycemic profile has been proposed for type 2 diabetic patients [198], [199]. With the aid of individuals who do not have diabetes, the differential equation with the delay model is used to create the non-linear model [200]. A popular "Uva/Padova Simulator" that had received FDA approval for the required clinical studies was also investigated. The parameters are derived for virtual type 1 diabetic patients [201]. Non-diabetic participants have experienced the OGTT [202]. To describe the approach using time monitoring, data from people with type 1 diabetes were gathered. For type 1 diabetic patients, the model is mathematically provided for BG value estimation in the PP mode [203], [204]. Using data from two days of clinical observations, the decision model for BG-insulin equilibrium over a more extended time is investigated [205]. For the goal of routine glucose monitoring, a meal identification algorithm was created for T1DM patients. The technique incorporates a mathematical model for bolus meals that delivers insulin and glucose [206]. To obtain the values for the inconsistent condition extent algorithm, diabetic and healthy patients were considered. An intelligent PID controller (iPID) was created for a person with type 1 diabetes after the food intake schedule was analyzed in the absence of food profiles to keep the balance of blood sugar level profiles [207], [208].

B. GLUCOSE CONTROLS CONSUMER PRODUCTS

When insulin secretion is not synchronized through the pancreas for proper glucose consumption. Then, glucose levels would be imbalanced. Anyone with type 1 diabetes must be prepared for insulin treatment. Patients who are going to be treated using insulin injections daily may also use an insulin pump. Insulin can also be administered via self-injection. Throughout the day, an insulin pump continuously administers short-acting insulin. The insulin pump has eliminated the need for long-acting insulin. A pump also helps to lower blood sugar levels and replaces the need for many daily injections and continuous insulin infusion. Numerous varieties of insulin doses are already in demand as commercial goods, namely insulin pumps from Animas, Medtronic, Roche, Tandem, and Omnipod. With regard to their enhanced features,

these insulin pumps are more sophisticated than one another. To have greater control over the glycemic profile, a state-of-the-art technique for measuring glucose is compared in Table 5.

VIII. MEASUREMENT AND CONTROLS OF GLUCOSE LEVELS: IOMT PERSPECTIVES

The requirement of an intelligent and automated healthcare system for diabetes patients in smart cities and smart villages to improve the quality of life [218], [219]. Continuous monitoring for emergency care, ambient intellect, & QoS for an appropriate PoC means are characteristics of smart healthcare systems [220], [221]. For diabetics, accurate and non-invasive glucose measurement is necessary, as well as the ability to save data utilizing IoMT for effective therapy [27]. Traditional glucose measurement techniques have limited capabilities and cannot support patients at remote locations. The diabetic individual would like the ability to store their data on a cloud server and regularly check their glycemic profile during the day. The point-of-care therapy for those with diabetes would be possible with regular monitoring, thanks to the smart healthcare system.

The virtual paradigm has enabled smart healthcare to connect persons to physicians at remote locations for quick remedy & treatment [218]. Using modern healthcare, consistent monitoring of critical problems provides appropriate treatment. Smart healthcare can improve service quality while lowering costs with the active assistance of remote healthcare solutions. The intelligent sensors would continually collect patient data and assist in storing it in a cloud data center. Additionally, it is helpful for data analysis and the simple information interchange between patients and doctors via mobile applications. The healthcare real-time system is successfully employed to apply intelligent algorithms to meet the many difficulties of the healthcare industry.




FIGURE 29: Diagnosis and management of blood sugar in a smart healthcare system.

The diabetic patient's diets may be planned with glucose management prescribed from a dietitian along with continuous glucose monitoring. To facilitate the CGM, the proposed device should be precise, affordable, and easy to use [29], [33]. The serum glucose measurement is always considered

TABLE 5: A comparison of a few state-of-the-art methods for measuring blood sugar

Technique	Sample	Summary	Findings
photoplethysmography (PPG) [86]	Finger	By using machine learning algorithms to extract PPG signal properties, it is possible to predict blood pressure, blood sugar, and systolic and diastolic values.	ML models applied where random forest technique has best prediction results as $R_{SBP}^2 = 0.91$, $R_{DBP}^2 = 0.89$ and $R_{BGL}^2 = 0.90$. CEG showed 80% values are in A zone
photoplethysmography (PPG) [165]	Finger	Multi-channel PPG along with Pulse Arrival Velocity (PAV) have been implemented for measurement	The dual-channel PPG with PAV's amplitude ratio produced the most accurate estimates of blood sugar level.
ATR spectroscopy [209]	oral mucosa inner lips	greater measurement repeatability and spectra about 1155 cm^{-1} for various blood glucose levels during fasting and before fasting were obtained using a multi-reflection prism and greater sensitivity.	the R^2 is above 0.7. The Sensitivity is very low, and all the measured values are in A region
Opto-chemical sensing [160]	Direct in the blood	The sensor is built on a basic optical fibre that may be introduced into the bloodstream using a catheter.	The device is evaluated to measure glucose in vivo using a non-human model. The sensor has a detection limit of 50.89 mg/dl and glucose sensitivity of 0.0354 a.u./mg.dl.
Millimeter, NIR wave [161]	Body object	Analysis of data from a multisensor network that combines the results of a random forest machine learning (ML) framework with information from supplementary sensors including mm-wave, near-infrared, and others.	The model predicts glucose with MARD, and RMSE values are high. The CECA shows the model's predictions fall within the boundaries of Zone A, which is a clinically acceptable range.
Optical Coherence Tomography [210]	Fingertip	It measures the optical rotation angle and depolarization index for aqueous glucose solutions with low and high scattering, respectively. Angle value increases as glucose value increases, although the depolarization index drops.	The correlation factor has a value of R^2 0.9101. The average deviation is around 0.027.
Contact lenses fluorescence [211]	Tears	Making a soft, intelligent contact lens that uses transparent, flexible nanostructures to fully integrate glucose sensors, wireless power transfer circuits, and display pixels to show sensing signals in real-time.	The use of smart and soft lenses would enable real-time wireless glucose monitoring in tears.
IMPS and mNIRS [212]	Left Hand and wrist Hand	IMPS and mNIRS use the indirect dielectric characteristics of the surrounding tissue around blood and the optical scattering characteristics of glucose itself in blood, respectively, the proposed IC can remove various systemic noises to enhance the glucose level estimation accuracy	8.3% - 15.0% mean absolute relative differences (mARD) using IMPS and 15.0–20.0% using mNIRS have been reported for the normal range of blood glucose levels. From CEG analysis, all of the measurement results are clinically acceptable of Zone A and 90% of total samples can be used for clinical treatment
Microwave Detection [213]	earlobe	The absorption spectrum of microwave signal helps to measure using two antennae. The sine wave of 500 MHz is for blood glucose measurement.	It can measure BG up to 500 mg/dl. It has a mean standard deviation (MAD) of 0.5226 while the minimum SD value is 0.04.
PPG [94]	Finger	AI was used with a cell phone camera to anticipate blood sugar levels. First, the incorrect data was isolated, and any type of training wasn't needed for the entire system.	The prototype could measure glucose in the normal range. The results show overall accuracy at the desired level
MEMS [214]	Finger	It is a minimally invasive procedure called e-Mosquito that uses a micro-actuator made of shape-memory alloy (SMA) to collect blood samples. It is regarded as the first wearable technology to automatically take blood from a patient and analyze their blood glucose levels.	The method provided good linear correlation in Conventional flow and the e-Mosquito prototype.
Transmission spectroscopy [215]	Saliva	Following completely absorbing the necessary quantity of saliva on the strip, the specimen would be transported to the detection zone via paper microfluidic activity, where a chemical reaction within GOx and salivary glucose begins to produce a pH change, leading to an adjustment in strip color that was tracked through the RGB detection on the small device that helps with recognizing glucose.	With very high sensitivity and a limit of detection (LOD) within a response time, the designed biosensor could detect glucose concentrations between appropriate range
VNIR [216]	Wrist	The biosensor enables the manipulation of pulsing of arterial blood from the wrist tissue. The VNIR spectroscopy was employed for reflected optical signals to estimate blood glucose.	The Rp value after averaging is approximately 0.86, while the standard prediction error is comparatively low.
MIR ATRS [209]	Inner lip mucosa	A novel optical fibre probe was developed utilizing a multi-reflection prism and ATR spectroscopy. Measurement repeatability was greater due to the prism's large, flat, wide contact surface and sensitivity with the number of reflections.	The experiment's findings show the presence of glucose in several spectra while fasting and after a glucose infusion. Peak absorption is defined by the calibration plot at 1155 cm^{-1} , with a glucose measurement error of less than 20%.
Short NIR [217]	Fingertip	A novel noninvasive glucometer with short NIR waves and light absorption and reflectance at specific wavelengths (940 and 1,300 nm) has been introduced. A unique, accurate ML-based method for glucose sensor calibration using calibrated and verified healthy, prediabetic, and diabetic samples has been reported.	Pearson's correlation coefficient (R) is 0.953, mARD is 7.32 %, and RMSE value is 11.96 mg/dl. All the values lines in Zone A (94%) and Zone B (6%).

more accurate than capillary analysis. Therefore, the continuous monitoring solution for quick serum glucose measurement is preferred for smart healthcare. The novel serum glucose monitoring-based device is a portable solution incorporating IoMT to continually record glucose data in the cloud. Patients located at remote locations might benefit by analyzing the data from medical professionals. The continuous glucose monitoring in smart healthcare is demonstrated in Fig. 29. Remote health monitoring is the need of diabetic patients and helps patients to long-term complications such as renal disease, CVD, sexual problems, vision issues, and airborne diseases. The other advantage of remote health monitoring is to provide a better diagnosis from available experts within a short duration. Patient monitoring from a remote location is also helpful for early diagnosis of diabetes and diet management activities. The patients will be aware of upcoming issues from their current lifestyles. The role of remote monitoring has become significant since the COVID-19 pandemic. Hence, patient remote monitoring is an integral component of smart healthcare.

A detailed description of a closed-loop system that displays glucose level monitoring and insulin discharge is shown (refer to Fig. 30 [2]). By Controlling diabetes using closed-loop automatic production of insulin, this virtual architecture can offer a superior method for evaluating insulin dosages. Such an integrated cloud framework may analyze diabetic




FIGURE 30: An IoMT framework-based closed-loop automated insulin secretion diabetes control system [2].

patients, treat them for blood glucose management in intelligent healthcare, and provide adequate healthcare in smart infrastructure with a few healthcare staff.

Medical device security and privacy concerns are the most important aspects of any IoT network. Because control operations often occur through wireless media, wearable device hardware security is extremely important. The control of the glucose monitoring device's security flaws is demonstrated in Fig. 31.

Due to linked health systems in an unstable and unsafe IoMT framework, device security is crucial [222]. Another




FIGURE 31: Our Long Time Goal of Security-Assured Non-invasive Sugar Level Measurement and Controlling in Our suggested iGLU

essential security component of smart healthcare is the reliability of usable medical data. All patient medical records are kept on a server. Thus, the safety of this information is crucial. To have safe monitoring and appropriate patient care regulated access with correct authentication is necessary.

IX. SHORTCOMINGS AND OPEN CHALLENGES OF EXISTING WORK

This section outlines the shortcomings and discusses some unresolved issues with controlling and measuring glucose levels.

A. LIMITATIONS OF CURRENT METHODS AND PRODUCTS

- 1) To estimate BG level, photoacoustic spectroscopy has been used. Real-time measurement and validation have not been done. In the laboratory, an artificial solution was made to monitor BG level. The prototype module with the laser and corresponding detector is expensive. The prototype needs much more space and does not offer a portable option. It is, therefore, not a better option for continuous glucose monitoring.
- 2) When monochromatic light interacts with a particular material, a nonlinear scattering process called Raman spectroscopy occurs. Raman spectroscopy can be used as a laboratory test, but it also takes up a lot more space. Therefore, the system built using this methodology won't work for regular glucose monitoring.
- 3) Data is also gathered using the eye retina for body glucose monitoring, which is one of the alternative non-invasive glucose detection approaches. Such a method is not always helpful for the measurement of glucose.
- 4) In the case of the bio-capacitance technique, a small change in the detector's position may impact the sensor's outcome, which reflects the change in BG level. The detector output may also be impacted by skin sweat, body temperature, and pressure on the sensor.
- 5) Impedance spectroscopy (IMPS) is used to detect glucose by connecting electrodes, influenced by skin. Since each person's saliva and sweat may differ, the accuracy of the glucose measurement won't be consistent.

As a result, this method is not accurate for measuring glucose in smart healthcare.

- 6) A PPG signal is employed to remove characteristics for predicting body glucose levels. However, the PPG can be an accurate body glucose-measuring method whose output value varies solely with blood volume. As a result, the presented method has not been able to detect the glucose molecule in the blood sample correctly.

B. OPEN CHALLENGES IN NON-INVASIVE BODY GLUCOSE MEASURING

The expansion of a non-invasive body glucose measurement device is analyzed with limitations. However, numerous unresolved issues that pose serious obstacles to accurate non-invasive glucose measurement have been explored. These difficulties have been identified in Fig. 32. Hypoglycemic patient monitoring is not possible without pricking for a long time as the non-invasive devices are required to be developed for precise measurement below 80 mg/dl [223].




FIGURE 32: Glucose-Level Measurement: Open Challenges.

- In the literature, factors that impact glucose measurement results include BP, temperature, and humidity level. However, these factors are not evaluated.
- Additionally, the affordable and portable continuous glucose technology has not been fully addressed.
- The accurate glucose measurement is still not possible at higher blood glucose range without the blood pricking method
- The development of a smart healthcare framework using the integration of a glucometer with the Internet of Med-

ical Things for continuous data logging to the cloud has not been visible in commercial devices [224].

- With the use of a cloud server, the decision model for automatically secreting insulin to monitor glucose values has to be explored.
- The privacy and security problems with the insulin and body glucose monitoring system have not been explored commercially.
- Continuously monitoring glucose with an insulin supply system requires an efficient power management system.
- The continuous glucose monitoring device would be portable and battery-operated so power management of the device needs to be addressed.

The non-invasive glucose measurement would be also an ideal choice for infants. Newborn babies tend to enter into hypoglycemia conditions during febrile seizure episodes. It is required to adopt the continuous glucose monitoring approach to check the glucose value during the high fever condition. request pricking of blood samples from newborn babies is irritating and panic. It is generally not recommended to prick multiple times as the body's blood level is comparatively lower. Continuous monitoring of other blood parameters for newborn babies would help to provide rapid diagnosis and treatment to avoid recurrent febrile seizure episodes. The non-invasive glucose measurement would be helpful for Type 1 diabetes patients for adjustment of their insulin dosage as per their current body glucose.

X. CONCLUSIONS

The study provides an overview of BG measurement, controlling mechanisms, and continuous monitoring strategies. Numerous methods have been described in the state of the art as proofs-of-concept, demonstrating a strong connection between device response and the reference value for blood glucose. Some techniques are not implemented for commercial purposes. Some approaches are neither accurate nor cost-effective solutions. The prior technologies are discussed with design strategies, observed issues, and measurement limitations. Due to the limitations and issues, advancements have also been discussed in terms of solutions. The main focus of the paper is to demonstrate various techniques with corresponding issues and solutions, along with advancements. Optical detection uses short NIR, which has been considered a future appliance or prototype device that should be more effective in various zones to support continuous health monitoring and viable solutions to reduce the shortcomings of all other techniques. Various methodologies may be utilized in the future for precise glucose monitoring. The consumer devices should be more effective in various zones to support continuous health monitoring. It must be used regularly as a portable device for real-time applications. Future devices should be low-cost and user-friendly for continuous health monitoring systems.

XI. FUTURE WORK

The future plan for a non-invasive glucose measurement device is highlighted in Fig. 33, which represents the future milestone with significant features. There is a requirement to develop a portable, durable, and user-friendly device so that it can be used on a large scale all over the world. The upcoming non-invasive devices should be able to check blood glucose for all age people precisely. The patient's data should be safe and confidential. The data can be accessed by patients and medical experts only. The upcoming device should be low-power and send a warning signal to patients after reaching at alarm level. These expected features are future milestones. An advanced IoMT framework integration for the device is necessary. This cutting-edge IoMT framework will connect the future measuring device with all nearby diabetes care centers for optimum care. Combining food intake and glucose measurement of particular can significantly provide the root cause and corresponding treatment for smart healthcare [225]. The future of technology expects reliable, portable, and user-friendly devices. The feature of borderline cross indication on the device should be explored in future measuring devices. Everyone will be aware to check their own BG level to analyze the body's proper function. A secured device with end-to-end user control and authentication is also necessary for future advancement. Physical unclonable function (PUF) based security framework is advantageous for IoMT devices [222], [226]. The effectiveness of a suitable healthcare cyber-physical system (H-CPS) with blockchain-based data and device management has to be considered [227], [228]. Glucose metres could be combined with wearables and food diaries to track individual reactions to various meals and activities. This would enable people to customize their diet and exercise routines for better performance and health.

ACKNOWLEDGMENT

The survey work has been carried out with joint collaboration of the authors. The authors would like to thank the MNIT Jaipur, where the literature work was started, and special thanks to other institutes for motivational support. This work was also in part supported by the Indo-Norwegian Collaboration in Autonomous Cyber-Physical Systems (INCAPS) project: 287918 of the International Partnerships for Excellent Education, Research and Innovation (INTPART) program from the Research Council of Norway.

REFERENCES

- [1] Diabetestalk, "What is the normal fasting blood sugar range for adults," 2018, last Accessed on 18 Jan 2021. [Online]. Available: <https://diabetestalk.net/blood-sugar/what-is-the-normal-fasting-blood-sugar-range-for-adults>
- [2] P. Jain, A. M. Joshi, and S. P. Mohanty, "iGLU 1.1: Towards a glucose-insulin model based closed loop IoMT framework for automatic insulin control of diabetic patients," in *2020 IEEE 6th World Forum on Internet of Things (WF-IoT)*. IEEE, 2020, pp. 1–6.
- [3] S. Garg, U. P. Shukla, and A. M. Joshi, "Bilstm calibrated iGLU with demographic data: Non-invasive glucose measurement device," in *2022 IEEE International Symposium on Smart Electronic Systems (iSES)*. IEEE, 2022, pp. 349–353.
- [4] A. M. Joshi, U. P. Shukla, and S. P. Mohanty, "Smart healthcare for diabetes during COVID-19," *IEEE Consumer Electronics Magazine*, vol. 10, no. 1, pp. 66–71, January 2020.
- [5] Amit M Joshi, U. P. Shukla, and S. P. Mohanty, "Smart healthcare for diabetes: A COVID-19 perspective," *arXiv preprint arXiv:2008.11153*, 2020.
- [6] A. M. Alsamman and H. Zayed, "The transcriptomic profiling of COVID-19 compared to SARS, MERS, Ebola, and H1N1," *bioRxiv*, 2020.
- [7] I. D. Federation, "IDF Diabetes Atlas - Diabetes is rising worldwide... and is set to rise even further," 2019, last Accessed on 21 March 2020. [Online]. Available: <https://diabetesatlas.org/en/sections/worldwide-toll-of-diabetes.html>
- [8] N. H. Cho, J. E. Shaw, S. Karuranga, Y. Huang, J. D. da Rocha Fernandes, A. W. Ohlrogge, and B. Malanda, "IDF diabetes atlas: Global estimates of diabetes prevalence for 2017 and projections for 2045," *Diabetes Research and Clinical Practice*, vol. 138, pp. 271–281, April 2018.
- [9] P. Saeedi, I. Petersohn, P. Salpea, B. Malanda, S. Karuranga, N. Unwin, S. Colagiuri, L. Guariguata, A. A. Motala, K. Ogurtsova et al., "Global and regional diabetes prevalence estimates for 2019 and projections for 2030 and 2045: Results from the international diabetes federation diabetes atlas," *Diabetes research and clinical practice*, vol. 157, p. 107843, 2019.
- [10] Clevelandclinic, "Diabetes Mellitus: An Overview," 2020, last Accessed on 18 Jan 2021. [Online]. Available: <https://my.clevelandclinic.org/health/diseases/7104-diabetes-mellitus-an-overview>
- [11] Drugs, "Type 1 diabetes mellitus," 2020, last Accessed on 18 Jan 2021. [Online]. Available: <https://www.drugs.com/health-guide/type-1-diabetes-mellitus.html>
- [12] L. M. Leontis and A. Hess-Fischl, "Diabetes mellitus: An overview," 2019, last Accessed on 18 Jan 2021. [Online]. Available: <https://www.endocrineweb.com/conditions/type-2-diabetes/type-2-diabetes-symptoms>
- [13] H. Agrawal, P. Jain, and A. M. Joshi, "Machine learning models for non-invasive glucose measurement: towards diabetes management in smart healthcare," *Health and Technology*, vol. 12, no. 5, pp. 955–970, 2022.
- [14] A. Pietrangelo, "What are the different types of diabetes?" 2018, last Accessed on 18 Jan 2021. [Online]. Available: <https://www.healthline.com/health/diabetes/types-of-diabetes>
- [15] M. J. Fowler, "Diabetes: Magnitude and mechanisms," *Clinical Diabetes*, vol. 28, no. 1, pp. 42–46, 2010. [Online]. Available: <https://clinical.diabetesjournals.org/content/28/1/42>
- [16] H. Yin, B. Mukadam, X. Dai, and N. Jha, "DiabDeep: Pervasive Diabetes Diagnosis based on Wearable Medical Sensors and Efficient Neural Networks," *IEEE Transactions on Emerging Topics in Computing*, pp. 1–1, 2019.
- [17] P. Zhang, "Global healthcare expenditure on diabetes for 2010 and 2030," *Diabetes Research and Clinical Practice*, 2011.
- [18] J. Venkataraman and B. Freer, "Feasibility of non-invasive blood glucose monitoring: In-vitro measurements and phantom models," in *2011 IEEE International Symposium on Antennas and Propagation (APSURSI)*, July 2011, pp. 603–606.
- [19] S. H. Wild, G. Roglic, A. Green, R. Sicree, and H. King, "Global prevalence of diabetes: Estimates for the year 2000 and projections for 2030," *Diabetes Care*, vol. 27, no. 10, pp. 2569–2569, 2004. [Online]. Available: <http://care.diabetesjournals.org/content/27/10/2569.2>
- [20] D. R. Whiting, L. Guariguata, C. Weil, and J. Shaw, "Idf diabetes atlas: Global estimates of the prevalence of diabetes for 2011 and 2030," *Diabetes Research and Clinical Practice*, vol. 94, no. 3, pp. 311 – 321, 2011.
- [21] P. H. Siegel, A. Tang, G. Virbila, Y. Kim, M. C. F. Chang, and V. Pikov, "Compact non-invasive millimeter-wave glucose sensor," in *2015 40th International Conference on Infrared, Millimeter, and Terahertz waves (IRMMW-THz)*, Aug 2015, pp. 1–3.
- [22] S. M. Alavi, M. Gourzi, A. Rouane, and M. Nadi, "An original method for non-invasive glucose measurement: preliminary results," in *2001 Conference Proceedings of the 23rd Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, vol. 4, 2001, pp. 3318–3320 vol.4.
- [23] X. Li and C. Li, "Study on the application of wavelet transform to non-invasive glucose concentration measurement by nirs," in *2015 Fifth International Conference on Instrumentation and Measurement, Computer, Communication and Control (IMCCC)*, Sept 2015, pp. 1294–1297.
- [24] P. P. Pai, P. K. Sanki, S. K. Sahoo, A. De, S. Bhattacharya, and S. Banerjee, "Cloud computing-based non-invasive glucose monitoring for diabetic




FIGURE 33: Our Long-Term Goals for Non-Invasive Glucose-Level Monitoring.

- care,” *IEEE Transactions on Circuits and Systems I: Regular Papers*, vol. PP, no. 99, pp. 1–14, 2017.
- [25] P. S. Reddy and K. Jyostna, “Development of smart insulin device for non invasive blood glucose level monitoring,” in *2017 IEEE 7th International Advance Computing Conference (IACC)*, Jan 2017, pp. 516–519.
- [26] A. M. Joshi, P. Jain, S. P. Mohanty, and N. Agrawal, “iGLU 2.0: A new wearable for accurate non-invasive continuous serum glucose measurement in IoMT framework,” *IEEE Transactions on Consumer Electronics*, no. 10.1109/TCE.2020.3011966, p. in Press, 2020.
- [27] S. P. Mohanty and E. Kougiannos, “Biosensors: A Tutorial Review,” *IEEE Potentials*, vol. 25, no. 2, pp. 35–40, March 2006.
- [28] N. A. Salam, W. H. M. Saad, Z. Manap, and F. Salehuddin, “The evolution of non-invasive blood glucose monitoring system for personal application,” *Journal of Telecommunication, Electronic and Computer Engineering*, vol. 8, pp. 59–65, 2016.
- [29] P. Jain, A. M. Joshi, and S. P. Mohanty, “iGLU 1.0: An Accurate Non-Invasive Near-Infrared Dual Short Wavelengths Spectroscopy based Glucometer for Smart Healthcare,” *arXiv Electrical Engineering and Systems Science*, no. arXiv:1911.04471, November 2019.
- [30] P. Jain, A. M. Joshi, N. Agrawal, and S. P. Mohanty, “iGLU 2.0: A New Non-invasive, Accurate Serum Glucometer for Smart Healthcare,” *arXiv Electrical Engineering and Systems Science*, vol. abs/2001.09182, 2020. [Online]. Available: <http://arxiv.org/abs/2001.09182>
- [31] Q. Liu, Y. Liu, F. Wu, X. Cao, Z. Li, M. Alharbi, A. N. Abbas, M. R. Amer, and C. Zhou, “Highly Sensitive and Wearable In2O3 Nanoribbon Transistor Biosensors with Integrated On-Chip Gate for Glucose Monitoring in Body Fluids,” *ACS Nano*, vol. 12, no. 2, pp. 1170–1178, 2018, pMID: 29338249.
- [32] P. H. Siegel, W. Dai, R. A. Kloner, M. Csete, and V. Pikov, “First millimeter-wave animal in vivo measurements of l-glucose and d-glucose: Further steps towards a non-invasive glucometer,” in *2016 41st International Conference on Infrared, Millimeter, and Terahertz waves (IRMMW-THz)*, Sept 2016, pp. 1–3.
- [33] P. Jain, A. M. Joshi, and S. P. Mohanty, “iGLU: An Intelligent Device for Accurate Non-Invasive Blood Glucose-Level Monitoring in Smart Healthcare,” *IEEE Consumer Electronics Magazine*, vol. 9, no. 1, pp. 35–42, January 2020.
- [34] N. M. Zhilo, P. A. Rudenko, and A. N. Zhigaylo, “Development of hardware-software test bench for optical non-invasive glucometer improvement,” in *2017 IEEE Conference of Russian Young Researchers in Electrical and Electronic Engineering (EIConRus)*, Feb 2017, pp. 89–90.
- [35] S. I. Gusev, A. A. Simonova, P. S. Demchenko, M. K. Khodzitsky, and O. P. Cherkasova, “Blood glucose concentration sensing using biological molecules relaxation times determination,” in *2017 IEEE International Symposium on Medical Measurements and Applications (MeMeA)*, May 2017, pp. 458–463.
- [36] M. W. Sari and M. Luthfi, “Design and analysis of non-invasive blood glucose levels monitoring,” in *2016 International Seminar on Application for Technology of Information and Communication (ISemantic)*, Aug 2016, pp. 134–137.
- [37] S. Lekha and M. Suchetha, “Non- invasive diabetes detection and classification using breath analysis,” in *2015 International Conference on Communications and Signal Processing (ICCSPP)*, April 2015, pp. 0955–0958.
- [38] J. Li, P. Koinkar, Y. Fuchiwaki, and M. Yasuzawa, “A fine pointed glucose oxidase immobilized electrode for low-invasive amperometric glucose monitoring,” *Biosensors and Bioelectronics*, vol. 86, pp. 90–94, 2016.
- [39] N. Demitri and A. M. Zoubir, “Measuring blood glucose concentrations in photometric glucometers requiring very small sample volumes,” *IEEE Transactions on Biomedical Engineering*, vol. 64, no. 1, pp. 28–39, 2017.
- [40] A. Sun, A. G. Venkatesh, and D. A. Hall, “A multi-technique reconfigurable electrochemical biosensor: Enabling personal health monitoring in mobile devices,” *IEEE Transactions on Biomedical Circuits and Systems*, vol. 10, no. 5, pp. 945–954, Oct 2016.
- [41] A. Gani, A. V. Gribok, Y. Lu, W. K. Ward, R. A. Vigersky, and J. Reifman, “Universal glucose models for predicting subcutaneous glucose concentration in humans,” *IEEE Transactions on Information Technology in Biomedicine*, vol. 14, no. 1, pp. 157–165, Jan 2010.
- [42] G. Wang, M. D. Poscente, S. S. Park, C. N. Andrews, O. Yadi-Pecht, and M. P. Mintchev, “Wearable microsystem for minimally invasive, pseudo-continuous blood glucose monitoring: The e-mosquito,” *IEEE Transactions on Biomedical Circuits and Systems*, vol. 11, no. 5, pp. 979–

- 987, Oct 2017.
- [43] M. M. Ahmadi and G. A. Jullien, "A wireless-implantable microsystem for continuous blood glucose monitoring," *IEEE Transactions on Biomedical Circuits and Systems*, vol. 3, no. 3, pp. 169–180, June 2009.
 - [44] I. Pagkalos, P. Herrero, C. Toumazou, and P. Georgiou, "Bio-inspired glucose control in diabetes based on an analogue implementation of a β -cell model," *IEEE Transactions on Biomedical Circuits and Systems*, vol. 8, no. 2, pp. 186–195, April 2014.
 - [45] T. Kossowski and R. Stasinski, "Robust ir attenuation measurement for non-invasive glucose level analysis," in *2016 International Conference on Systems, Signals and Image Processing (IWSSIP)*, May 2016, pp. 1–4.
 - [46] L. P. Pavlovich and D. Y. Mynziak, "Noninvasive method for blood glucose measuring and monitoring," in *2013 IEEE XXXIII International Scientific Conference Electronics and Nanotechnology (ELNANO)*, April 2013, pp. 255–257.
 - [47] Y. Liu, W. Li, T. Zheng, and W. K. Ling, "Overviews the methods of non-invasive blood glucose measurement," in *2016 IEEE International Conference on Consumer Electronics-China (ICCE-China)*, Dec 2016, pp. 1–2.
 - [48] N. K. Sharma and S. Singh, "Designing a non invasive blood glucose measurement sensor," in *2012 IEEE 7th International Conference on Industrial and Information Systems (ICIIS)*, Aug 2012, pp. 1–3.
 - [49] X. Zhao, Q. Zheng, and Z. M. Yang, "Two types of photonic crystals applied to glucose sensor," in *2016 IEEE International Nanoelectronics Conference (INEC)*, May 2016, pp. 1–2.
 - [50] Y. Tanaka, C. Purtil, T. Tajima, M. Seyama, and H. Koizumi, "Sensitivity improvement on cw dual-wavelength photoacoustic spectroscopy using acoustic resonant mode for noninvasive glucose monitor," in *2016 IEEE SENSORS*, Oct 2016, pp. 1–3.
 - [51] I. Gouzouasis, H. Cano-Garcia, I. Sotiriou, S. Saha, G. Palikaras, P. Kosmas, and E. Kallos, "Detection of varying glucose concentrations in water solutions using a prototype biomedical device for millimeter-wave non-invasive glucose sensing," in *2016 10th European Conference on Antennas and Propagation (EuCAP)*, April 2016, pp. 1–4.
 - [52] Y. Nikawa and D. Someya, "Non-invasive measurement of blood sugar level by millimeter waves," in *2001 IEEE MTT-S International Microwave Symposium Digest (Cat. No.01CH37157)*, vol. 1, May 2001, pp. 171–174 vol.1.
 - [53] J. Shao, F. Yang, F. Xia, Q. Zhang, and Y. Chen, "A novel miniature spiral sensor for non-invasive blood glucose monitoring," in *2016 10th European Conference on Antennas and Propagation (EuCAP)*, April 2016, pp. 1–2.
 - [54] P. H. Siegel, Y. Lee, and V. Pivov, "Millimeter-wave non-invasive monitoring of glucose in anesthetized rats," in *2014 39th International Conference on Infrared, Millimeter, and Terahertz waves (IRMMW-THz)*, Sept 2014, pp. 1–2.
 - [55] D. Wang, "An improved integration sensor of non-invasive blood glucose," in *The 7th IEEE/International Conference on Advanced Infocomm Technology*, Nov 2014, pp. 70–75.
 - [56] N. Bayasi, H. Saleh, B. Mohammad, and M. Ismail, "The revolution of glucose monitoring methods and systems: A survey," in *2013 IEEE 20th International Conference on Electronics, Circuits, and Systems (ICECS)*, Dec 2013, pp. 92–93.
 - [57] R. Agrawal, N. Sharma, M. Rathore, V. Gupta, S. Jain, V. Agarwal, and S. Goyal, "Noninvasive method for glucose level estimation by saliva," *J Diabetes Metab*, vol. 4, no. 5, pp. 2–5, 2013.
 - [58] N. Demitri and A. M. Zoubir, "Measuring blood glucose concentrations in photometric glucometers requiring very small sample volumes," *IEEE Transactions on Biomedical Engineering*, vol. 64, no. 1, pp. 28–39, 2016.
 - [59] M. Shokrehkodaie and S. Quinones, "Review of Non-invasive Glucose Sensing Techniques: Optical, Electrical and Breath Acetone," *Sensors*, vol. 20, no. 5, p. 1251, 2020.
 - [60] S. Delbeck, T. Vahlsing, S. Leonhardt, G. S. G. and H. M. Heise, "Non-invasive monitoring of blood glucose using optical methods for skin spectroscopy-opportunities and recent advances," *Anal Bioanal Chem.*, vol. 411, no. 1, pp. 63–77, 2019.
 - [61] N. K. Madzhi, S. A. Shamsuddin, and M. F. Abdullah, "Comparative investigation using gaas(950nm), gaaias (940nm) and ingaasp (1450nm) sensors for development of non-invasive optical blood glucose measurement system," in *2014 IEEE International Conference on Smart Instrumentation, Measurement and Applications (ICSIMA)*, Nov 2014, pp. 1–6.
 - [62] N. A. M. Aziz, N. Arsad, P. S. Menon, A. R. Laili, M. H. Laili, and A. A. A. Halim, "Analysis of difference light sources for non-invasive aqueous glucose detection," in *2014 IEEE 5th International Conference on Photonics (ICP)*, Sept 2014, pp. 150–152.
 - [63] S. Tommasone, "Infrared spectroscopy: An overview," 2018, last Accessed on 18 Jan 2021. [Online]. Available: <https://www.azolifesciences.com/article/Infrared-Spectroscopy-An-Overview.aspx>
 - [64] A. A. Muley and R. B. Ghongade, "Design and simulate an antenna for aqueous glucose measurement," in *2014 Annual IEEE India Conference (INDICON)*, Dec 2014, pp. 1–6.
 - [65] K. U. Menon, D. Hemachandran, and A. T. Kunnath, "Voltage intensity based non-invasive blood glucose monitoring," in *2013 Fourth International Conference on Computing, Communications and Networking Technologies (ICCCNT)*. IEEE, 2013, pp. 1–5.
 - [66] J. L. Lai, S. Y. Huang, R. S. Lin, and S. C. Tsai, "Design a non-invasive near-infrared led blood glucose sensor," in *2016 International Conference on Applied System Innovation (ICASI)*, May 2016, pp. 1–4.
 - [67] M. Tamilselvi and G. Ramkumar, "Non-invasive tracking and monitoring glucose content using near infrared spectroscopy," in *2015 IEEE International Conference on Computational Intelligence and Computing Research (ICCIC)*, Dec 2015, pp. 1–3.
 - [68] K. Lawand, M. Parihar, and S. N. Patil, "Design and development of infrared led based non invasive blood glucometer," in *2015 Annual IEEE India Conference (INDICON)*, Dec 2015, pp. 1–6.
 - [69] J. Yadav, A. Rani, V. Singh, and B. M. Murari, "Near-infrared led based non-invasive blood glucose sensor," in *2014 International Conference on Signal Processing and Integrated Networks (SPIN)*, Feb 2014, pp. 591–594.
 - [70] M. T. B. Z. Abidin, M. K. R. Rosli, S. A. Shamsuddin, N. K. Madzhi, and M. F. Abdullah, "Initial quantitative comparison of 940nm and 950nm infrared sensor performance for measuring glucose non-invasively," in *2013 IEEE International Conference on Smart Instrumentation, Measurement and Applications (ICSIMA)*, Nov 2013, pp. 1–6.
 - [71] Y. Nikawa and T. Michiyama, "Non-invasive measurement of blood-sugar level by reflection of millimeter-waves," in *2006 Asia-Pacific Microwave Conference*, Dec 2006, pp. 47–50.
 - [72] Y. Zhao, L. Li, B. Han, W. Zheng, X. Li, and Y.-n. Zhang, "Ultra-sensitive optical fiber glucose sensor based on in-fiber mzi with vernier effect," *Journal of Lightwave Technology*, pp. 1–8, 2023.
 - [73] M. Goodarzi and W. Saeys, "Selection of the most informative near infrared spectroscopy wavebands for continuous glucose monitoring in human serum," *Talanta*, vol. 146, pp. 155–165, 2016.
 - [74] S. Sharma, M. Goodarzi, L. Wynants, H. Ramon, and W. Saeys, "Efficient use of pure component and interferent spectra in multivariate calibration," *Analytica chimica acta*, vol. 778, pp. 15–23, 2013.
 - [75] Y. Uwadaira, N. Adachi, A. Ikehata, and S. Kawano, "Factors affecting the accuracy of non-invasive blood glucose measurement by short-wavelength near infrared spectroscopy in the determination of the glycaemic index of foods," *Journal of Near Infrared Spectroscopy*, vol. 18, no. 5, pp. 291–300, 2010.
 - [76] S. Haxha and J. Jhoja, "Optical based noninvasive glucose monitoring sensor prototype," *IEEE Photonics Journal*, vol. 8, no. 6, pp. 1–11, 2016.
 - [77] W. Zhang, R. Liu, W. Zhang, H. Jia, and K. Xu, "Discussion on the validity of nir spectral data in non-invasive blood glucose sensing," *Biomedical optics express*, vol. 4, no. 6, pp. 789–802, 2013.
 - [78] M. Golic, K. Walsh, and P. Lawson, "Short-wavelength near-infrared spectra of sucrose, glucose, and fructose with respect to sugar concentration and temperature," *Applied spectroscopy*, vol. 57, no. 2, pp. 139–145, 2003.
 - [79] K. Song, U. Ha, S. Park, J. Bae, and H. J. Yoo, "An impedance and multi-wavelength near-infrared spectroscopy ic for non-invasive blood glucose estimation," *IEEE Journal of Solid-State Circuits*, vol. 50, no. 4, pp. 1025–1037, April 2015.
 - [80] S. Ramasahayam, K. S. Haindavi, and S. R. Chowdhury, "Noninvasive estimation of blood glucose concentration using near infrared optodes," in *Sensing Technology: Current Status and Future Trends IV*. Springer, 2015, pp. 67–82.
 - [81] A. Heller, "Integrated medical feedback systems for drug delivery," *AIChe journal*, vol. 51, no. 4, pp. 1054–1066, 2005.
 - [82] P. P. Pai, P. K. Sanki, A. De, and S. Banerjee, "NIR photoacoustic spectroscopy for non-invasive glucose measurement," in *2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*. IEEE, 2015, pp. 7978–7981.
 - [83] F. Shaikh, N. Haworth, R. Wells, J. Bishop, S. K. Chatterjee, S. Banerjee, and S. Laha, "Compact instrumentation for accurate detection and mea-

- surement of glucose concentration using photoacoustic spectroscopy," *IEEE Access*, vol. 10, pp. 31 885–31 895, 2022.
- [84] A. M. Joshi, K. Divya, H. Chhajed, and R. S. Kamal, "FPGA implementation of multivariate support vector regression for non-invasive blood glucose estimation using IoT framework," in *IoT Applications for Healthcare Systems*. Springer, 2022, pp. 77–90.
 - [85] P. Jain, R. Maddila, and A. M. Joshi, "A precise non-invasive blood glucose measurement system using NIR spectroscopy and Huber's regression model," *Optical and Quantum Electronics*, vol. 51, no. 2, p. 51, 2019.
 - [86] E. Monte-Moreno, "Non-invasive estimate of blood glucose and blood pressure from a photoplethysmograph by means of machine learning techniques," *Artificial intelligence in medicine*, vol. 53, no. 2, pp. 127–138, 2011.
 - [87] S. Habbu, M. Dale, and R. Ghongade, "Estimation of blood glucose by non-invasive method using photoplethysmography," *Sādhanā*, vol. 44, no. 6, p. 135, 2019.
 - [88] H. Ali, F. Bensali, and F. Jaber, "Novel approach to non-invasive blood glucose monitoring based on transmittance and refraction of visible laser light," *IEEE Access*, vol. 5, pp. 9163–9174, 2017.
 - [89] C. Vrančić, A. Fomichova, N. Gretz, C. Herrmann, S. Neudecker, A. Pucci, and W. Petrich, "Continuous glucose monitoring by means of mid-infrared transmission laser spectroscopy in vitro," *Analyst*, vol. 136, pp. 1192–1198, 2011. [Online]. Available: <http://dx.doi.org/10.1039/C0AN00537A>
 - [90] B. Paul, M. P. Manuel, and Z. C. Alex, "Design and development of non invasive glucose measurement system," in *2012 1st International Symposium on Physics and Technology of Sensors (ISPTS-1)*. IEEE, 2012, pp. 43–46.
 - [91] L. A. Philip, K. Rajasekaran, and E. S. J. Jothi, "Continuous monitoring of blood glucose using photoplethysmograph signal," in *2017 International Conference on Innovations in Electrical, Electronics, Instrumentation and Media Technology (ICEEIMT)*. IEEE, 2017, pp. 187–191.
 - [92] H. Karimipour, H. T. Shandiz, and E. Zahedi, "Diabetic diagnose test based on ppg signal and identification system," *Journal of Biomedical Science and Engineering*, vol. 2, no. 06, p. 465, 2009.
 - [93] F. R. G. Cruz, C. C. Paglinawan, C. N. V. Catindig, J. C. B. Lamchek, D. D. C. Almiranez, and A. F. Sanchez, "Application of reflectance mode photoplethysmography for non-invasive monitoring of blood glucose level with moving average filter," in *Proceedings of the 2019 9th International Conference on Biomedical Engineering and Technology*. ACM, 2019, pp. 22–26.
 - [94] Y. Zhang, Y. Zhang, S. A. Siddiqui, and A. Kos, "Non-invasive blood-glucose estimation using smartphone ppg signals and subspace knn classifier," *Elektrotehnički Vestnik*, vol. 86, no. 1/2, pp. 68–74, 2019.
 - [95] Y. Yamakoshi, K. Matsumura, T. Yamakoshi, J. Lee, P. Rolfe, Y. Kato, K. Shimizu, and K.-i. Yamakoshi, "Side-scattered finger-photoplethysmography: experimental investigations toward practical non-invasive measurement of blood glucose," *Journal of biomedical optics*, vol. 22, no. 6, p. 067001, 2017.
 - [96] O. Olarte, W. V. Moer, K. Barbé, Y. V. Ingelgem, and A. Hubin, "Influence of the type and position of the sensor on the precision of impedance glucose measurements," in *2013 IEEE International Instrumentation and Measurement Technology Conference (I2MTC)*, May 2013, pp. 1750–1754.
 - [97] S. K. Dhar, P. Biswas, and S. Chakraborty, "Dc impedance of human blood using eis: An approach to non-invasive blood glucose measurement," in *2013 International Conference on Informatics, Electronics and Vision (ICIEV)*, May 2013, pp. 1–6.
 - [98] Y. Khawam, M. Ali, H. Shazada, S. Kanan, and H. Nashash, "Non-invasive blood glucose measurement using transmission spectroscopy," in *2013 1st International Conference on Communications, Signal Processing, and their Applications (ICCSPA)*, Feb 2013, pp. 1–4.
 - [99] M. N. Anas and P. K. Lim, "A bio-impedance approach," in *2013 IEEE International Conference on Smart Instrumentation, Measurement and Applications (ICSIMA)*, Nov 2013, pp. 1–5.
 - [100] C. E. F. Amaral and B. Wolf, "Effects of glucose in blood and skin impedance spectroscopy," in *AFRICON 2007*, Sept 2007, pp. 1–7.
 - [101] P. Jain and A. M. Joshi, "Low leakage and high cmrr cmos differential amplifier for biomedical application," *Analog Integrated Circuits and Signal Processing*, pp. 1–15, 2017.
 - [102] B. Paul, M. P. Manuel, and Z. C. Alex, "Design and development of non invasive glucose measurement system," in *2012 1st International Symposium on Physics and Technology of Sensors (ISPTS-1)*, March 2012, pp. 43–46.
 - [103] M. Hofmann, M. Bloss, R. Weigel, G. Fischer, and D. Kissinger, "Non-invasive glucose monitoring using open electromagnetic waveguides," in *2012 42nd European Microwave Conference*, Oct 2012, pp. 546–549.
 - [104] Y. Liu, M. Xia, Z. Nie, J. Li, Y. Zeng, and L. Wang, "In vivo wearable non-invasive glucose monitoring based on dielectric spectroscopy," in *2016 IEEE 13th International Conference on Signal Processing (ICSP)*, Nov 2016, pp. 1388–1391.
 - [105] M. Yamaguchi, M. Mitsumori, and Y. Kano, "Noninvasively measuring blood glucose using saliva," *IEEE Engineering in Medicine and Biology Magazine*, vol. 17, no. 3, pp. 59–63, May 1998.
 - [106] M. S. Prasad, R. Chen, Y. Li, D. Rekha, D. Li, H. Ni, and N. Y. Sreedhar, "Polypyrrole supported with copper nanoparticles modified alkali anodized steel electrode for probing of glucose in real samples," *IEEE Sensors Journal*, vol. 18, no. 13, pp. 5203–5212, July 2018.
 - [107] G. Yoon, K. J. Jeon, A. K. Amerov, Y.-J. Kim, D. Y. Hwang, J. B. Kim, and H. S. Kim, "Non-invasive monitoring of blood glucose," in *Lasers and Electro-Optics, 1999. CLEO/Pacific Rim '99. The Pacific Rim Conference on*, vol. 4, Aug 1999, pp. 1233–1234 vol.4.
 - [108] Y. Yamakoshi, M. Ogawa, T. Yamakoshi, M. Satoh, M. Nogawa, S. Tanaka, T. Tamura, P. Rolfe, and K. Yamakoshi, "A new non-invasive method for measuring blood glucose using instantaneous differential near infrared spectrophotometry," in *2007 29th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, Aug 2007, pp. 2964–2967.
 - [109] G. Yoon, A. K. Amerov, K. J. Jeon, J. B. Kim, and Y.-J. Kim, "Optical measurement of glucose levels in scattering media," in *Proceedings of the 20th Annual International Conference of the IEEE Engineering in Medicine and Biology Society. Vol.20 Biomedical Engineering Towards the Year 2000 and Beyond (Cat. No.98CH36286)*, vol. 4, Oct 1998, pp. 1897–1899 vol.4.
 - [110] H. Ishizawa, A. Muro, T. Takano, K. Honda, and H. Kanai, "Non-invasive blood glucose measurement based on atr infrared spectroscopy," in *2008 SICE Annual Conference*, Aug 2008, pp. 321–324.
 - [111] T. Harada, K. Yamamoto, M. Kondo, K. Gesho, and I. Ishimaru, "Spectroscopy optical coherence tomography of biomedical tissue," in *SICE Annual Conference 2007*, Sept 2007, pp. 3056–3059.
 - [112] A. Popov, A. Bykov, S. Toppari, M. Kinnunen, A. Priezhev, and R. Myllylä, "Glucose sensing in flowing blood and intralipid by laser pulse time-of-flight and optical coherence tomography techniques," *IEEE Journal of Selected Topics in Quantum Electronics*, vol. 18, pp. 1335–1342, 07 2012.
 - [113] J. Y. Sim, C. G. Ahn, E. Jeong, and B. K. Kim, "Photoacoustic spectroscopy that uses a resonant characteristic of a microphone for in vitro measurements of glucose concentration," in *2016 38th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, Aug 2016, pp. 4861–4864.
 - [114] P. P. Pai, P. K. Sanki, and S. Banerjee, "A photoacoustics based continuous non-invasive blood glucose monitoring system," in *2015 IEEE International Symposium on Medical Measurements and Applications (MeMeA) Proceedings*, May 2015, pp. 106–111.
 - [115] P. P. Pai, P. K. Sanki, A. De, and S. Banerjee, "Nir photoacoustic spectroscopy for non-invasive glucose measurement," in *2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, Aug 2015, pp. 7978–7981.
 - [116] Y. Tanaka, Y. Higuchi, and S. Camou, "Noninvasive measurement of aqueous glucose solution at physiologically relevant blood concentration levels with differential continuous-wave laser photoacoustic technique," in *2015 IEEE SENSORS*, Nov 2015, pp. 1–4.
 - [117] L. Xiaoli and L. Chengwei, "Research on glucose concentration sensing with single wavelength laser," in *2015 12th IEEE International Conference on Electronic Measurement Instruments (ICEMI)*, vol. 03, July 2015, pp. 1547–1551.
 - [118] H. A. A. Naam, M. O. Idrees, A. Awad, O. S. Abdalsalam, and F. Mohamed, "Non invasive blood glucose measurement based on photoacoustic spectroscopy," in *2015 International Conference on Computing, Control, Networking, Electronics and Embedded Systems Engineering (ICCNEE)*, Sept 2015, pp. 1–4.
 - [119] S. Camou, Y. Ueno, and E. Tamechika, "New cw-photoacoustic-based protocol for noninvasive and selective determination of aqueous glucose level: A potential alternative towards noninvasive blood sugar sensing," in *2011 IEEE SENSORS Proceedings*, Oct 2011, pp. 798–801.
 - [120] N. Wadamori, R. Shinohara, and Y. Ishihara, "Photoacoustic depth profiling of a skin model for non-invasive glucose measurement," in *2008 30th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, Aug 2008, pp. 5644–5647.

- [121] S. Koyama, Y. Miyauchi, T. Horiguchi, and H. Ishizawa, "Non-invasive measurement of blood glucose of diabetic based on ir spectroscopy," in *Proceedings of SICE Annual Conference 2010*, Aug 2010, pp. 3425–3426.
- [122] P. Domachuk, M. Hunter, R. Batorsky, M. Cronin-Golomb, F. G. Omenetto, A. Wang, A. K. George, and J. C. Knight, "A path for non-invasive glucose detection using mid-ir supercontinuum," in *2008 Conference on Lasers and Electro-Optics and 2008 Conference on Quantum Electronics and Laser Science*, May 2008, pp. 1–2.
- [123] R. Periyasamy and S. Anand, "A study on non-invasive blood glucose estimation- an approach using capacitance measurement technique," in *2016 International Conference on Signal Processing, Communication, Power and Embedded System (SCOPES)*, Oct 2016, pp. 847–850.
- [124] T. Yilmaz, R. Foster, and Y. Hao, "Towards accurate dielectric property retrieval of biological tissues for blood glucose monitoring," *IEEE Transactions on Microwave Theory and Techniques*, vol. 62, no. 12, pp. 3193–3204, Dec 2014.
- [125] M. Gourzi, A. Rouane, M. B. McHugh, R. Guelaz, and M. Nadi, "New biosensor for non-invasive glucose concentration measurement," in *Proceedings of IEEE Sensors 2003 (IEEE Cat. No.03CH37498)*, vol. 2, Oct 2003, pp. 1343–1347 Vol.2.
- [126] V. Turgul and I. Kale, "On the accuracy of complex permittivity model of glucose/water solutions for non-invasive microwave blood glucose sensing," in *2015 E-Health and Bioengineering Conference (EHB)*, Nov 2015, pp. 1–4.
- [127] D. Li, D. Yang, J. Yang, Y. Lin, Y. Sun, H. Yu, and K. Xu, "Glucose affinity measurement by surface plasmon resonance with borate polymer binding," *Sensors and Actuators A: Physical*, vol. 222, pp. 58 – 66, 2015. [Online]. Available: <http://www.sciencedirect.com/science/article/pii/S0924424714004737>
- [128] R. Kaul and U. P. Khot, "Design of microstrip antennas for glucometer application," in *2016 IEEE International Conference on Advances in Electronics, Communication and Computer Technology (ICAECCT)*, Dec 2016, pp. 352–357.
- [129] V. Turgul and I. Kale, "Influence of fingerprints and finger positioning on accuracy of rf blood glucose measurement from fingertips," *Electronics Letters*, vol. 53, no. 4, pp. 218–220, 2017.
- [130] H. Cano-Garcia, I. Gouzouasis, I. Sotiriou, S. Saha, G. Palikaras, P. Kosmas, and E. Kallos, "Reflection and transmission measurements using 60 ghz patch antennas in the presence of animal tissue for non-invasive glucose sensing," in *2016 10th European Conference on Antennas and Propagation (EuCAP)*, April 2016, pp. 1–3.
- [131] S. Saha, I. Sotiriou, I. Gouzouasis, H. Cano-Garcia, G. Palikaras, P. Kosmas, and E. Kallos, "Evaluation of the sensitivity of transmission measurements at millimeter waves using patch antennas for non-invasive glucose sensing," in *2016 10th European Conference on Antennas and Propagation (EuCAP)*, April 2016, pp. 1–4.
- [132] V. Turgul and I. Kale, "A novel pressure sensing circuit for non-invasive rf/microwave blood glucose sensors," in *2016 16th Mediterranean Microwave Symposium (MMS)*, Nov 2016, pp. 1–4.
- [133] M. S. Ali, N. J. Shoumy, S. Khatun, L. M. Kamarudin, and V. Vijayaravswari, "Non-invasive blood glucose measurement performance analysis through uwb imaging," in *2016 3rd International Conference on Electronic Design (ICED)*, Aug 2016, pp. 513–516.
- [134] H. Choi, J. Nylon, S. Luzio, J. Beutler, and A. Porch, "Design of continuous non-invasive blood glucose monitoring sensor based on a microwave split ring resonator," in *2014 IEEE MTT-S International Microwave Workshop Series on RF and Wireless Technologies for Biomedical and Healthcare Applications (IMWS-Bio2014)*, Dec 2014, pp. 1–3.
- [135] V. Turgul and I. Kale, "Simulating the effects of skin thickness and fingerprints to highlight problems with non-invasive rf blood glucose sensing from fingertips," *IEEE Sensors Journal*, vol. 17, no. 22, pp. 7553–7560, Nov 2017.
- [136] S. Sindhuja and E. Kanniga, "Flexible antenna sensor in thumb spica splint for noninvasive monitoring of fluctuating blood glucose levels," *IEEE Sensors Journal*, vol. 23, no. 1, pp. 544–551, 2023.
- [137] Y. Miyauchi, T. Horiguchi, H. Ishizawa, S. i. Tezuka, and H. Hara, "Blood glucose level measurement by confocal reflection photodetection system," in *SICE Annual Conference 2011*, Sept 2011, pp. 2686–2689.
- [138] D. W. Kim, H. S. Kim, D. H. Lee, and H. C. Kim, "Importance of skin resistance in the reverse iontophoresis-based noninvasive glucose monitoring system," in *The 26th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, vol. 1, Sept 2004, pp. 2434–2437.
- [139] M. Hofmann, T. Fersch, R. Weigel, G. Fischer, and D. Kissinger, "A novel approach to non-invasive blood glucose measurement based on rf transmission," in *2011 IEEE International Symposium on Medical Measurements and Applications*, May 2011, pp. 39–42.
- [140] K. Mitsubayashi, "Novel biosensing devices for medical applications soft contact-lens sensors for monitoring tear sugar," in *2014 International Conference on Simulation of Semiconductor Processes and Devices (SISPAD)*, Sept 2014, pp. 349–352.
- [141] B. D. Cameron and G. L. Cote, "Polarimetric glucose sensing in aqueous humor utilizing digital closed-loop control," in *Proceedings of 18th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, vol. 1, Oct 1996, pp. 204–205 vol.1.
- [142] R. J. Buford, E. C. Green, and M. J. McClung, "A microwave frequency sensor for non-invasive blood-glucose measurement," in *2008 IEEE Sensors Applications Symposium*, Feb 2008, pp. 4–7.
- [143] S. Lee, V. Nayak, J. Dodds, M. Pishkou, and N. B. Smith, "Glucose measurements with sensors and ultrasound," *Ultrasound in Medicine and Biology*, vol. 31, no. 7, pp. 971–977, 2005.
- [144] O. K. Cho, Y. O. Kim, H. Mitsumaki, and K. Kuwa, "Noninvasive Measurement of Glucose by Metabolic Heat Conformation Method," *Clinical Chemistry*, vol. 50, no. 10, pp. 1894–1898, 10 2004. [Online]. Available: <https://doi.org/10.1137/clinchem.2004.036954>
- [145] A. B. Blodgett, R. K. Kothinti, I. Kamysenko, D. H. Petering, S. Kumar, and N. M. Tabatabai, "A fluorescence method for measurement of glucose transport in kidney cells," *Diabetes Technology & Therapeutics*, vol. 13, no. 7, pp. 743–751, 2011, PMID: 21510766.
- [146] A. K. Amerov, Y. Sun, G. W. Small, and M. A. Arnold, "Kromoscopic measurement of glucose in the first overtone region of the near-infrared spectrum," in *Optical Diagnostics and Sensing of Biological Fluids and Glucose and Cholesterol Monitoring II*, A. V. Priezhev and G. L. Cote, Eds., vol. 4624, International Society for Optics and Photonics. SPIE, 2002, pp. 11 – 19. [Online]. Available: <https://doi.org/10.1117/12.468318>
- [147] R. Zhang, S. Liu, H. Jin, Y. Luo, Z. Zheng, F. Gao, and Y. Zheng, "Noninvasive electromagnetic wave sensing of glucose," *Sensors*, vol. 19, no. 5, p. 1151, 2019.
- [148] J. Malik, S. Kim, J. M. Seo, Y. M. Cho, and F. Bien, "Minimally invasive implant type electromagnetic biosensor for continuous glucose monitoring system: In vivo evaluation," *IEEE Transactions on Biomedical Engineering*, vol. 70, no. 3, pp. 1000–1011, 2023.
- [149] P. Bertemes-Filho, R. Weinert, T. Barato, and T. Baratto de Albuquerque, "Detection of glucose by using impedance spectroscopy," in *Proc. International Conference on Electrical Bio-Impedance*, 06 2016.
- [150] P. Arpaia, F. Mancino, and N. Moccaldi, "A reproducible bioimpedance transducer for insulin noninvasive measurement," *IEEE Transactions on Instrumentation and Measurement*, vol. 72, pp. 1–11, 2023.
- [151] R. O. Potts, J. A. Tamada, and M. J. Tierney, "Glucose monitoring by reverse iontophoresis," *Diabetes/Metabolism Research and Reviews*, vol. 18, no. S1, pp. S49–S53, 2002. [Online]. Available: <https://onlinelibrary.wiley.com/doi/abs/10.1002/dmrr.210>
- [152] R. Rao and S. Nanda, "Sonophoresis: recent advancements and future trends," *Journal of Pharmacy and Pharmacology*, vol. 61, no. 6, pp. 689–705, 2009. [Online]. Available: <https://onlinelibrary.wiley.com/doi/abs/10.1211/jpp.61.06.0001>
- [153] O. Amir, D. Weinstein, S. Zilberman, M. Less, D. Perl-Treves, H. Primack, A. Weinstein, E. Gabis, B. Fikhte, and A. Karasik, "Continuous noninvasive glucose monitoring technology based on "occlusion spectroscopy"," *Journal of Diabetes Science and Technology*, vol. 1, no. 4, pp. 463–469, 2007, PMID: 19885108.
- [154] B. M. Jensen, P. Bjerring, J. S. Christiansen, and H. ørskov, "Glucose content in human skin: relationship with blood glucose levels," *Scandinavian Journal of Clinical and Laboratory Investigation*, vol. 55, no. 5, pp. 427–432, 1995.
- [155] T. Kossowski and R. Stasiński, "Multi-wavelength analysis of substances levels in human blood," in *2017 International Conference on Systems, Signals and Image Processing (IWSSIP)*, May 2017, pp. 1–4.
- [156] A. Ficorella, A. D'Amico, M. Santonic, G. Pennazza, S. Grasso, and A. Zompanti, "A multi-frequency system for glucose detection with optical sensors," in *2015 XVIII AISEM Annual Conference*, Feb 2015, pp. 1–3.
- [157] K. Song, U. Ha, S. Park, and H.-J. Yoo, "An impedance and multi-wavelength near-infrared spectroscopy ic for non-invasive blood glucose estimation," in *2014 Symposium on VLSI Circuits Digest of Technical Papers*, June 2014, pp. 1–2.

- [158] E. L. Litinskaia, N. A. Bazaev, K. V. Pozhar, and V. M. Grinvald, "Methods for improving accuracy of non-invasive blood glucose detection via optical glucometer," in *2017 IEEE Conference of Russian Young Researchers in Electrical and Electronic Engineering (EIConRus)*, Feb 2017, pp. 47–49.
- [159] A. M. Joshi, P. Jain, and S. P. Mohanty, "iGLU 3.0: A secure noninvasive glucometer and automatic insulin delivery system in IoMT," *IEEE Transactions on Consumer Electronics*, vol. 68, no. 1, pp. 14–22, 2022.
- [160] S. Kundu, S. Tabassum, R. A. Kumar, E. Dale Abel, and R. Kumar, "Plasmonic optical fiber based continuous in-vivo glucose monitoring for icu/ccu setup," *IEEE Transactions on NanoBioscience*, pp. 1–1, 2023.
- [161] Y. Sun, H. Cano-Garcia, E. Kallos, F. O'Brien, A. Akintonde, D.-E. Motei, O. Ancu, R. W. A. Mackenzie, and P. Kosmas, "Random forest analysis of combined millimeter-wave and near-infrared sensing for noninvasive glucose detection," *IEEE Sensors Journal*, vol. 23, no. 17, pp. 20 294–20 309, 2023.
- [162] Y. Wei, J. Liu, L. Hu, B. W.-K. Ling, and Q. Liu, "Time frequency analysis-based averaging and fusion of features for wearable non-invasive blood glucose estimation," *IEEE Transactions on Consumer Electronics*, vol. 69, no. 3, pp. 510–521, 2023.
- [163] A. K. Singh and S. K. Jha, "Fabrication and validation of a handheld non-invasive, optical biosensor for self-monitoring of glucose using saliva," *IEEE Sensors Journal*, vol. 19, no. 18, pp. 8332–8339, Sep. 2019.
- [164] M. Nakagawa, K. Oiwa, Y. Nanai, K. Nagumo, and A. Nozawa, "Feature extraction for estimating acute blood glucose level variation from multiwavelength facial images," *IEEE Sensors Journal*, vol. 23, no. 17, pp. 20 247–20 257, 2023.
- [165] P.-L. Lee, K.-W. Wang, and C.-Y. Hsiao, "A non-invasive blood glucose estimation system using dual-channel ppgs and pulse-arrival velocity," *IEEE Sensors Journal*, pp. 1–1, 2023.
- [166] P. P. Pai, A. De, and S. Banerjee, "Accuracy enhancement for noninvasive glucose estimation using dual-wavelength photoacoustic measurements and kernel-based calibration," *IEEE Transactions on Instrumentation and Measurement*, vol. 67, no. 1, pp. 126–136, 2018.
- [167] T. Dai and A. Adler, "In vivo blood characterization from bioimpedance spectroscopy of blood pooling," *IEEE Transactions on Instrumentation and Measurement*, vol. 58, no. 11, p. 3831, 2009.
- [168] R. D. Beach, R. W. Conlan, M. C. Godwin, and F. Moussy, "Towards a miniature implantable in vivo telemetry monitoring system dynamically configurable as a potentiostat or galvanostat for two- and three-electrode biosensors," *IEEE Transactions on Instrumentation and Measurement*, vol. 54, no. 1, pp. 61–72, Feb 2005.
- [169] P. Jain, S. Pancholi, and A. M. Joshi, "An IoMT based non-invasive precise blood glucose measurement system," in *2019 IEEE International Symposium on Smart Electronic Systems (iSES)(Formerly iNiS)*, IEEE, 2019, pp. 111–116.
- [170] B. A. Malik, A. Naqash, and G. M. Bhat, "Backpropagation artificial neural network for determination of glucose concentration from near-infrared spectra," in *2016 International Conference on Advances in Computing, Communications and Informatics (ICACCI)*, Sept 2016, pp. 2688–2691.
- [171] D. Yotha, C. Pidthalek, S. Yimman, and S. Niramitmahapanya, "Design and construction of the hypoglycemia monitor wireless system for diabetic," in *2016 9th Biomedical Engineering International Conference (BMEiCON)*, Dec 2016, pp. 1–4.
- [172] S. Sarangi, P. P. Pai, P. K. Sanki, and S. Banerjee, "Comparative analysis of golay code based excitation and coherent averaging for non-invasive glucose monitoring system," in *2014 IEEE 27th International Symposium on Computer-Based Medical Systems*, May 2014, pp. 485–486.
- [173] S. R. Naqvi, N. Z. Azeemi, A. Hameed, R. Baddar, and T. Rasool, "Improving accuracy of non-invasive glucose monitoring through non-local data denoising," in *2008 Cairo International Biomedical Engineering Conference*, Dec 2008, pp. 1–4.
- [174] Y. Yamakoshi, M. Ogawa, T. Yamakoshi, T. Tamura, and K. i. Yamakoshi, "Multivariate regression and discriminant calibration models for a novel optical non-invasive blood glucose measurement method named pulse glucometry," in *2009 Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, Sept 2009, pp. 126–129.
- [175] C. Z. Ming, P. Raveendran, and P. S. Chew, "A comparison analysis between partial least squares and neural network in non-invasive blood glucose concentration monitoring system," in *2009 International Conference on Biomedical and Pharmaceutical Engineering*, Dec 2009, pp. 1–4.
- [176] S. Sarangi, P. P. Pai, P. K. Sanki, and S. Banerjee, "Comparative analysis of golay code based excitation and coherent averaging for non-invasive glucose monitoring system," in *2014 IEEE 27th International Symposium on Computer-Based Medical Systems*, May 2014, pp. 485–486.
- [177] H. M. Heise, "Technology for non-invasive monitoring of glucose," in *Proceedings of 18th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, vol. 5, Oct 1996, pp. 2159–2161 vol.5.
- [178] P. P. Pai, S. Bhattacharya, and S. Banerjee, "Regularized least squares regression for calibration of a photoacoustic spectroscopy based non-invasive glucose monitoring system," in *2015 IEEE International Ultrasonics Symposium (IUS)*, Oct 2015, pp. 1–4.
- [179] M. Stemmann, F. Stahl, J. Lallemand, E. Renard, and R. Johansson, "Sensor calibration models for a non-invasive blood glucose measurement sensor," in *2010 Annual International Conference of the IEEE Engineering in Medicine and Biology*, Aug 2010, pp. 4979–4982.
- [180] D. K. Rollins, K. Kotz, and C. Stiehl, "Non-invasive glucose monitoring from measured inputs," in *UKACC International Conference on Control 2010*, Sept 2010, pp. 1–5.
- [181] B. A. Malik, "Determination of glucose concentration from near infrared spectra using least square support vector machine," in *2015 International Conference on Industrial Instrumentation and Control (ICIC)*, May 2015, pp. 475–478.
- [182] M. Ogawa, Y. Yamakoshi, M. Satoh, M. Nogawa, T. Yamakoshi, S. Tanaka, P. Rolfe, T. Tamura, and K. i. Yamakoshi, "Support vector machines as multivariate calibration model for prediction of blood glucose concentration using a new non-invasive optical method named pulse glucometry," in *2007 29th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, Aug 2007, pp. 4561–4563.
- [183] Z. T. Dag, E. Koklukaya, F. Temurtas, H. M. Saraoglu, and S. Altikat, "Detection of the blood glucose and haemoglobin a1c with palm perspiration by using artificial neural networks," in *2011 7th International Conference on Electrical and Electronics Engineering (ELECO)*, Dec 2011, pp. II–302–II–305.
- [184] O. Olarte, W. V. Moer, K. Barbe, Y. V. Ingelgem, and A. Hubin, "Using random phase multisines to perform non-invasive glucose measurements," in *2011 IEEE International Symposium on Medical Measurements and Applications*, May 2011, pp. 300–304.
- [185] K. Song, U. Ha, S. Park, and H.-J. Yoo, "An impedance and multi-wavelength near-infrared spectroscopy ic for non-invasive blood glucose estimation," in *2014 Symposium on VLSI Circuits Digest of Technical Papers*, June 2014, pp. 1–2.
- [186] R. Baghbani, M. A. Rad, and A. Pourziad, "Microwave sensor for non-invasive glucose measurements design and implementation of a novel linear," *IET Wireless Sensor Systems*, vol. 5, no. 2, pp. 51–57, 2015.
- [187] T. Zhu, L. Kuang, J. Daniels, P. Herrero, K. Li, and P. Georgiou, "IoMT-enabled real-time blood glucose prediction with deep learning and edge computing," *IEEE Internet of Things Journal*, vol. 10, no. 5, pp. 3706–3719, 2023.
- [188] P. Jain, A. Joshi, and S. Mohanty, "iGLU 4.0: A continuous glucose monitoring and balancing paradigm with physiological parameters," *arXiv preprint arXiv:2308.11952*, 2023.
- [189] J. S. Parab, R. S. Gad, and G. M. Naik, "Influence of pca components on glucose prediction using non-invasive technique," in *2016 International Conference on Advances in Electrical, Electronic and Systems Engineering (ICAEEs)*, Nov 2016, pp. 473–476.
- [190] T. R. J. C. Lekha and C. S. Kumar, "Nir spectroscopic algorithm development for glucose detection," in *2015 International Conference on Innovations in Information, Embedded and Communication Systems (ICIIECS)*, March 2015, pp. 1–6.
- [191] O. Olarte, W. V. Moer, K. Barbe, S. Verguts, Y. V. Ingelgem, and A. Hubin, "Using the best linear approximation as a first step to a new non-invasive glucose measurement," in *2012 IEEE International Instrumentation and Measurement Technology Conference Proceedings*, May 2012, pp. 2747–2751.
- [192] W. L. Clarke, "The original Clarke error grid analysis (EGA)," *Diabetes Technology & Therapeutics*, vol. 7, no. 5, pp. 776–779, 2005.
- [193] C. Fernandez, "Needle-free diabetes care: 7 devices that painlessly measure blood glucose," *Labitech*, vol. 23, 2018.
- [194] M. F. Schemmann and T. O'brien, "Blood glucose sensor," Apr. 11 2013, uS Patent App. 13/646,721.
- [195] A. Gal, I. Harman-Boehm, E. Naidis, Y. Mayzel, and L. Trieman, "Validity of glucotrack®), a non-invasive glucose monitor, for variety of diabetics," *Age*, vol. 1, no. 295, p. 61, 2011.
- [196] D. Huber, L. Falco-Jonasson, M. Talary, F. Dewarrat, A. Caduff, W. Stahel, and N. Stadler, "Multi-sensor data fusion for non-invasive continuous

- glucose monitoring,” in *2007 10th International Conference on Information Fusion*, July 2007, pp. 1–10.
- [197] P. Jain, A. Joshi, and S. Mohanty, “iGLU 4.0: Intelligent non-invasive glucose measurement and its control with physiological parameters,” *Springer Nature Computer Science (SN-CS)*, vol. XX, no. YY, p. NN, Accepted on 15 Feb 2024.
- [198] V. W. Bolie, “Coefficients of normal blood glucose regulation,” *Journal of applied physiology*, vol. 16, no. 5, pp. 783–788, 1961.
- [199] R. N. Bergman, L. S. Phillips, and C. Cobelli, “Physiologic evaluation of factors controlling glucose tolerance in man: measurement of insulin sensitivity and beta-cell glucose sensitivity from the response to intravenous glucose,” *The Journal of clinical investigation*, vol. 68, no. 6, pp. 1456–1467, 1981.
- [200] A. De Gaetano and O. Arino, “Mathematical modelling of the intravenous glucose tolerance test,” *Journal of mathematical biology*, vol. 40, no. 2, pp. 136–168, 2000.
- [201] C. Cobelli, C. Dalla Man, G. Sparacino, L. Magni, G. De Nicolao, and B. P. Kovatchev, “Diabetes: models, signals, and control,” *IEEE reviews in biomedical engineering*, vol. 2, pp. 54–96, 2009.
- [202] R. Hovorka, F. Shojaee-Moradie, P. V. Carroll, L. J. Chassin, I. J. Gowrie, N. C. Jackson, R. S. Tudor, A. M. Umpleby, and R. H. Jones, “Partitioning glucose distribution/transport, disposal, and endogenous production during ivgtt,” *American Journal of Physiology-Endocrinology and Metabolism*, vol. 282, no. 5, pp. E992–E1007, 2002.
- [203] A. Haidar, M. E. Wilinska, J. A. Graveston, and R. Hovorka, “Stochastic virtual population of subjects with type 1 diabetes for the assessment of closed-loop glucose controllers,” *IEEE Transactions on Biomedical Engineering*, vol. 60, no. 12, pp. 3524–3533, 2013.
- [204] H. Kirchsteiger, G. C. Estrada, S. Pölzer, E. Renard, and L. del Re, “Estimating interval process models for type 1 diabetes for robust control design,” *IFAC Proc. Volumes*, vol. 44, no. 1, pp. 11 761–11 766, 2011.
- [205] N. Magdelaine, L. Chaillous, I. Guilhem, J.-Y. Poirier, M. Krempf, C. H. Moog, and E. Le Carpentier, “A long-term model of the glucose–insulin dynamics of type 1 diabetes,” *IEEE Transactions on Biomedical Engineering*, vol. 62, no. 6, pp. 1546–1552, 2015.
- [206] K. Turksoy, S. Samadi, J. Feng, E. Littlejohn, L. Quinn, and A. Cinar, “Meal detection in patients with type 1 diabetes: a new module for the multivariable adaptive artificial pancreas control system,” *IEEE journal of biomedical and health informatics*, vol. 20, no. 1, pp. 47–54, 2015.
- [207] J. Xie and Q. Wang, “A variable state dimension approach to meal detection and meal size estimation: in silico evaluation through basal-bolus insulin therapy for type 1 diabetes,” *IEEE Transactions on Biomedical Engineering*, vol. 64, no. 6, pp. 1249–1260, 2016.
- [208] T. MohammadRidha, M. Ait-Ahmed, L. Chaillous, M. Krempf, I. Guilhem, J.-Y. Poirier, and C. H. Moog, “Model free ipid control for glycemia regulation of type-1 diabetes,” *IEEE Transactions on Biomedical Engineering*, vol. 65, no. 1, pp. 199–206, 2017.
- [209] S. Kino, S. Omori, T. Katagiri, and Y. Matsuura, “Hollow optical-fiber based infrared spectroscopy for measurement of blood glucose level by using multi-reflection prism,” *Biomedical optics express*, vol. 7, no. 2, pp. 701–708, 2016.
- [210] T.-L. Chen, Y.-L. Lo, C.-C. Liao, and Q.-H. Phan, “Noninvasive measurement of glucose concentration on human fingertip by optical coherence tomography,” *Journal of biomedical optics*, vol. 23, no. 4, p. 047001, 2018.
- [211] J. Park, J. Kim, S.-Y. Kim, W. H. Cheong, J. Jang, Y.-G. Park, K. Na, Y.-T. Kim, J. H. Heo, C. Y. Lee et al., “Soft, smart contact lenses with integrations of wireless circuits, glucose sensors, and displays,” *Science Advances*, vol. 4, no. 1, p. eaap9841, 2018.
- [212] K. Song, U. Ha, S. Park, J. Bae, and H.-J. Yoo, “An impedance and multi-wavelength near-infrared spectroscopy ic for non-invasive blood glucose estimation,” *IEEE Journal of solid-state circuits*, vol. 50, no. 4, pp. 1025–1037, 2015.
- [213] Q. Li, X. Xiao, and T. Kikkawa, “Absorption spectrum for non-invasive blood glucose concentration detection by microwave signals,” *Journal of Electromagnetic Waves and Applications*, vol. 33, no. 9, pp. 1093–1106, 2019.
- [214] G. Wang, M. D. Poscente, S. S. Park, C. N. Andrews, O. Yadid-Pecht, and M. P. Mintchev, “Wearable microsystem for minimally invasive, pseudo-continuous blood glucose monitoring: The e-mosquito,” *IEEE Transactions on Biomedical Circuits and Systems*, vol. 11, no. 5, pp. 979–987, 2017.
- [215] A. K. Singh and S. K. Jha, “Fabrication and validation of a handheld non-invasive, optical biosensor for self-monitoring of glucose using saliva,” *IEEE Sensors Journal*, vol. 19, no. 18, pp. 8332–8339, 2019.
- [216] V. P. Rachim and W.-Y. Chung, “Wearable-band type visible-near infrared optical biosensor for non-invasive blood glucose monitoring,” *Sensors and Actuators B: Chemical*, vol. 286, pp. 173–180, 2019.
- [217] P. Jain, A. M. Joshi, and S. P. Mohanty, “iGLU: An intelligent device for accurate noninvasive blood glucose-level monitoring in smart health-care,” *IEEE Consumer Electronics Magazine*, vol. 9, no. 1, pp. 35–42, 2019.
- [218] S. P. Mohanty, U. Choppali, and E. Kougianos, “Everything you wanted to know about smart cities: The Internet of things is the backbone,” *IEEE Consumer Electronics Magazine*, vol. 5, no. 3, pp. 60–70, July 2016.
- [219] P. Chanak and I. Banerjee, “Internet of things-enabled smart villages: Recent advances and challenges,” *IEEE Consumer Electronics Magazine*, pp. 1–1, 2020.
- [220] C. P. Antonopoulos, G. Keramidas, N. S. Voros, M. Huebner, F. Schwiagelshohn, D. Goehring, M. Dagioglou, G. Stavrinou, S. Konstantopoulos, and V. Karkaletsis, “Toward an ict-based service oriented health care paradigm,” *IEEE Consumer Electronics Magazine*, vol. 9, no. 4, pp. 77–82, 2020.
- [221] M. Aazam, S. Zeadally, and K. A. Harras, “Health fog for smart health-care,” *IEEE Consumer Electronics Magazine*, vol. 9, no. 2, pp. 96–102, 2020.
- [222] A. M. Joshi, P. Jain, and S. P. Mohanty, “Secure-iGLU: A secure device for noninvasive glucose measurement and automatic insulin delivery in IoMT framework,” in *2020 IEEE Computer Society Annual Symposium on VLSI (ISVLSI)*. IEEE, 2020, pp. 440–445.
- [223] P. Jain, A. M. Joshi, and S. Mohanty, “Everything you wanted to know about noninvasive glucose measurement and control,” *arXiv preprint arXiv:2101.08996*, 2021.
- [224] A. M. Joshi, P. Jain, and S. P. Mohanty, “Everything you wanted to know about continuous glucose monitoring,” *IEEE Consumer Electronics Magazine*, vol. 10, no. 6, pp. 61–66, 2021.
- [225] L. Rachakonda, S. P. Mohanty, and E. Kougianos, “iLog: An intelligent device for automatic food intake monitoring and stress detection in the IoMT,” *IEEE Transactions on Consumer Electronics*, vol. 66, no. 2, pp. 115–124, 2020.
- [226] V. P. Yanambaka, S. P. Mohanty, E. Kougianos, and D. Puthal, “PMsec: Physical unclonable function-based robust and lightweight authentication in the internet of medical things,” *IEEE Transactions on Consumer Electronics*, vol. 65, no. 3, pp. 388–397, 2019.
- [227] L. Rachakonda, A. K. Bapatla, S. P. Mohanty, and E. Kougianos, “SaY-oPillow: Blockchain-integrated privacy-assured IoMT framework for stress management considering sleeping habits,” *IEEE Transactions on Consumer Electronics*, pp. 1–1, 2020.
- [228] S. P. Mohanty, V. P. Yanambaka, E. Kougianos, and D. Puthal, “PUFchain: A hardware-assisted blockchain for sustainable simultaneous device and data security in the internet of everything (IoE),” *IEEE Consumer Electronics Magazine*, vol. 9, no. 2, pp. 8–16, 2020.



PRATEEK JAIN (Member, IEEE) earned his B.E. degree in Electronics Engineering from Jiwaji University, India in 2010 and his Master's degree from ITM University Gwalior. He has completed his PhD from MNIT Jaipur (Raj.) In 2020. He was awarded from MHRD fellowship. He is presently an Assistant Professor in the E & I Department of Nirma University Ahmedabad, India. He published several research papers in peer-reviewed Journals and Conferences. He is also a reviewer of reputed

Journals and conferences. He got 3 research seed grants at the university level. He is a member of IEEE, IEI and SPIE. His research interest is in Real-time System Design for Biomedical Applications, Analog Circuits, Digital VLSI Design and Machine Learning for Computation etc.



AMIT M. JOSHI (Senior Member, IEEE) received the M.Tech. and Ph.D. degrees from NIT, Surat, in 2009 and 2015, respectively. He has been an Associate Professor with the Malaviya National Institute of Technology, Jaipur (MNIT Jaipur), since December 2023. He has published more than 90 research articles in excellent peer-reviewed international journals/conferences and also has published six book chapters. He has a total of 875 Google Scholar citations, i10 index is 29, and H-index is

16. His research interests include biomedical signal processing, smart healthcare, VLSI DSP systems, and embedded system design. He is a member of IETE. He served as a Technical Program Committee Member for IEEE Conferences, such as iSES, ICCE, ISVLSI, and VDAT. He also received the honor of UGC Travel Fellowship, the Award of SERB DST Travel Grant, and CSIR Fellowship. He has attended well-known IEEE Conferences, such as TENCON-16, TENCON-17, ISCAS-18, and MENACOMM-19 across the world. He served as a reviewer for technical journals, such as IEEE TRANSACTIONS/IEEE ACCESS, Springer, and Elsevier.



SARAJU P. MOHANTY (Senior Member, IEEE) received the bachelor's degree (Honors) in electrical engineering from the Orissa University of Agriculture and Technology, Bhubaneswar, in 1995, the master's degree in Systems Science and Automation from the Indian Institute of Science, Bengaluru, in 1999, and the Ph.D. degree in Computer Science and Engineering from the University of South Florida, Tampa, in 2003. He is a Professor with the University of North Texas.

His research is in "Smart Electronic Systems" which has been funded by National Science Foundations (NSF), Semiconductor Research Corporation (SRC), U.S. Air Force, IUSSTF, and Mission Innovation. He has authored 500 research articles, 5 books, and 10 granted and pending patents. His Google Scholar h-index is 57 and i10-index is 243 with 13,000 citations. He is regarded as a visionary researcher on Smart Cities technology in which his research deals with security and energy aware, and AI/ML-integrated smart components. He introduced the Secure Digital Camera (SDC) in 2004 with built-in security features designed using Hardware Assisted Security (HAS) or Security by Design (SbD) principle. He is widely credited as the designer for the first digital watermarking chip in 2004 and first the low-power digital watermarking chip in 2006. He is a recipient of 19 best paper awards, Fulbright Specialist Award in 2021, IEEE Consumer Electronics Society Outstanding Service Award in 2020, the IEEE-CS-TCVLSI Distinguished Leadership Award in 2018, and the PROSE Award for Best Textbook in Physical Sciences and Mathematics category in 2016. He has delivered 24 keynotes and served on 14 panels at various International Conferences.

He has been serving on the editorial board of several peer-reviewed international transactions/journals, including IEEE Transactions on Big Data (TBD), IEEE Transactions on Computer-Aided Design of Integrated Circuits and Systems (TCAD), IEEE Transactions on Consumer Electronics (TCE), and ACM Journal on Emerging Technologies in Computing Systems (JETC). He has been the Editor-in-Chief (EiC) of the IEEE Consumer Electronics Magazine (MCE) during 2016-2021. He served as the Chair of Technical Committee on Very Large Scale Integration (TCVLSI), IEEE Computer Society (IEEE-CS) during 2014-2018 and on the Board of Governors of the IEEE Consumer Electronics Society during 2019-2021. He serves on the steering, organizing, and program committees of several international conferences. He is the steering committee chair/vice-chair for the IEEE International Symposium on Smart Electronic Systems (IEEE-iSES), the IEEE-CS Symposium on VLSI (ISVLSI), and the OITS International Conference on Information Technology (OCIT). He has mentored 3 post-doctoral researchers, and supervised 15 Ph.D. dissertations, 27 M.S. theses, and 27 undergraduate projects.



LINGA REDDY CENKERAMADDI (Senior Member, IEEE) received the master's degree in electrical engineering from the Indian Institute of Technology Delhi (IIT Delhi), New Delhi, India, in 2004, and the Ph.D. degree in electrical engineering from the Norwegian University of Science and Technology (NTNU), Trondheim, Norway, in 2011. He was with Texas Instruments on mixed signal circuit design before joining the Ph.D. program with NTNU. After finishing his Ph.D. degree,

he worked on radiation imaging for an atmosphere-space interaction monitor (ASIM mission to the International Space Station) with the University of Bergen, Bergen, Norway, from 2010 to 2012. He is currently a Leader of the Autonomous and Cyber-Physical Systems (ACPS) Research Group and a Professor with the University of Agder, Grimstad, Norway. He is a Principal Investigator and a Co-Principal Investigator of many research grants from the Norwegian Research Council. He is also quite active in medical imaging. His research interests include the Internet of Things (IoT), cyber-physical systems, autonomous systems, robotics and automation involving advanced sensor systems, computer vision, thermal imaging, LiDAR imaging, radar imaging, wireless sensor networks, smart electronic systems, advanced machine learning techniques, connected autonomous systems, including drones/unmanned aerial vehicles (UAVs), unmanned ground vehicles (UGVs), unmanned underwater systems (UUSs), 5G- (and beyond) enabled autonomous vehicles, and socio-technical systems, such as urban transportation systems, smart agriculture, and smart cities. He has coauthored more than 160 research publications that have been published in prestigious international journals and standard conferences in the above research areas. He is a member of ACM. Several of his master's students received the Best Master's Thesis Award in information and communication technology (ICT). He serves as a reviewer for several reputed international conferences and IEEE journals. He is a member of the editorial boards of various international journals and technical program committees of several IEEE conferences.

...